

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture **SDG No.:** P4397
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4397 **SAS No.:** P4397

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164298BS							
Arsenic	ug/L	4000	3630		91	80 - 120	P
Barium	ug/L	1000	924		92	80 - 120	P
Cadmium	ug/L	1000	885		88	80 - 120	P
Chromium	ug/L	2000	2070		104	80 - 120	P
Lead	ug/L	5000	4410		88	80 - 120	P
Selenium	ug/L	10000	8700		87	80 - 120	P
Silver	ug/L	380	360		95	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: P4397

Contract: PORT06

Lab Code: CHEM

Case No.: P4397

SAS No.: P4397

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164306BS Mercury	ug/L	4.0	3.52		88	80 - 120	CV