

Continuing Calibration Report for SequenceID : FC101424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067429.D

Aliphatic C9-C12	8508762.000	60.000	3.175	6.457	141812.700	128354.488	-10.485
Aliphatic C12-C16	5652181.000	40.000	6.458	9.850	141304.525	130957.677	-7.901
Aliphatic C16-C21	8170666.000	60.000	9.851	13.209	136177.767	131476.102	-3.576
Aliphatic C21-C28	10286818.000	80.000	13.210	16.865	128585.225	126823.664	-1.389
Aliphatic C28-C40	13123649.000	120.000	16.866	21.722	109363.742	105215.848	-3.942
Aliphatic EPH	45742076.000	360.000	3.175	21.722	127061.322	121110.937	-4.913