

Continuing Calibration Report for SequenceID : FC101424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067440.D

Aliphatic C9-C12	8663443.000	60.000	3.175	6.457	144390.717	128354.488	-12.494
Aliphatic C12-C16	5701770.000	40.000	6.458	9.850	142544.250	130957.677	-8.848
Aliphatic C16-C21	8254144.000	60.000	9.851	13.209	137569.067	131476.102	-4.634
Aliphatic C21-C28	10417464.000	80.000	13.210	16.865	130218.300	126823.664	-2.677
Aliphatic C28-C40	13488399.000	120.000	16.866	21.722	112403.325	105215.848	-6.831
Aliphatic EPH	46525220.000	360.000	3.175	21.722	129236.722	121110.937	-6.709