

Continuing Calibration Report for SequenceID : FD101424AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048504.D

Aromatic C10-C12	7355966.000	40.000	4.099	5.817	183899.150	180984.840	-1.610
Aromatic C12-C16	11202398.000	60.000	5.818	8.424	186706.633	189995.313	1.731
Aromatic C16-C21	13872207.000	80.000	8.425	12.687	173402.588	178728.230	2.980
Aromatic C21-C36	29243252.000	180.000	12.688	18.094	162462.511	165360.750	1.753
Aromatic EPH	61673823.000	360.000	4.099	18.094	171316.175	174173.072	1.640