

Continuing Calibration Report for SequenceID : FD101424AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048515.D

Aromatic C10-C12	7503437.000	40.000	4.099	5.817	187585.925	180984.840	-3.647
Aromatic C12-C16	11393796.000	60.000	5.818	8.424	189896.600	189995.313	0.052
Aromatic C16-C21	14125711.000	80.000	8.425	12.687	176571.388	178728.230	1.207
Aromatic C21-C36	29284546.000	180.000	12.688	18.094	162691.922	165360.750	1.614
Aromatic EPH	62307490.000	360.000	4.099	18.094	173076.361	174173.072	0.630