

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** P4397 **SAS No :** P4397 **SDG No:** P4397
Sample No : PB164109BSD **Datafile:** FC067432.D
Client ID : PB164109BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10	5.88	59		0.747 (40-140)	50
Aliphatic C12-C16	6.7	5.36	80		1.3 (40-140)	50
Aliphatic C16-C21	10	8.33	83		1.5 (40-140)	50
Aliphatic C21-C28	13.3	10.9	82		1.8 (40-140)	50
Aliphatic C28-C40	20.0	19.5	98		0.0025 (40-140)	50