

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

Continuing Calib Date: 10/14/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 09:11 Initial Calibration Time(s): 16:30 23:46

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.92       | 5.92      | 5.82      | 6.02  | 0.00       |
| Aroclor-1016-2       | (2) | 5.94       | 5.94      | 5.84      | 6.04  | 0.00       |
| Aroclor-1016-3       | (3) | 6.01       | 6.01      | 5.91      | 6.11  | 0.00       |
| Aroclor-1016-4       | (4) | 6.11       | 6.10      | 6.00      | 6.20  | -0.01      |
| Aroclor-1016-5       | (5) | 6.40       | 6.40      | 6.30      | 6.50  | 0.00       |
| Aroclor-1260-1       | (1) | 7.53       | 7.52      | 7.42      | 7.62  | -0.01      |
| Aroclor-1260-2       | (2) | 7.78       | 7.78      | 7.68      | 7.88  | 0.00       |
| Aroclor-1260-3       | (3) | 8.14       | 8.14      | 8.04      | 8.24  | 0.00       |
| Aroclor-1260-4       | (4) | 8.38       | 8.38      | 8.28      | 8.48  | 0.00       |
| Aroclor-1260-5       | (5) | 8.72       | 8.71      | 8.61      | 8.81  | -0.01      |
| Tetrachloro-m-xylene |     | 4.76       | 4.76      | 4.66      | 4.86  | 0.00       |
| Decachlorobiphenyl   |     | 10.68      | 10.67     | 10.57     | 10.77 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

Continuing Calib Date: 10/14/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 09:11 Initial Calibration Time(s): 16:30 23:46

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 5.16       | 5.16      | 5.06      | 5.26 | 0.00       |
| Aroclor-1016-2       | (2) | 5.18       | 5.18      | 5.08      | 5.28 | 0.00       |
| Aroclor-1016-3       | (3) | 5.36       | 5.36      | 5.26      | 5.46 | 0.00       |
| Aroclor-1016-4       | (4) | 5.40       | 5.40      | 5.30      | 5.50 | 0.00       |
| Aroclor-1016-5       | (5) | 5.62       | 5.62      | 5.52      | 5.72 | 0.00       |
| Aroclor-1260-1       | (1) | 6.66       | 6.66      | 6.56      | 6.76 | 0.00       |
| Aroclor-1260-2       | (2) | 6.85       | 6.85      | 6.75      | 6.95 | 0.00       |
| Aroclor-1260-3       | (3) | 7.01       | 7.01      | 6.91      | 7.11 | 0.00       |
| Aroclor-1260-4       | (4) | 7.48       | 7.48      | 7.38      | 7.58 | 0.00       |
| Aroclor-1260-5       | (5) | 7.72       | 7.72      | 7.62      | 7.82 | 0.00       |
| Tetrachloro-m-xylene |     | 4.05       | 4.05      | 3.95      | 4.15 | 0.00       |
| Decachlorobiphenyl   |     | 9.22       | 9.22      | 9.12      | 9.32 | 0.00       |