

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

Continuing Calib Date: 10/15/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 23:50 Initial Calibration Time(s): 16:30 23:46

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.92	5.92	5.82	6.02	0.00
Aroclor-1248-2 (2)	6.19	6.19	6.09	6.29	0.00
Aroclor-1248-3 (3)	6.40	6.40	6.30	6.50	0.00
Aroclor-1248-4 (4)	6.80	6.80	6.70	6.90	0.00
Aroclor-1248-5 (5)	6.84	6.84	6.74	6.94	0.00
Tetrachloro-m-xylene	4.75	4.75	4.65	4.85	0.00
Decachlorobiphenyl	10.67	10.67	10.57	10.77	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

Continuing Calib Date: 10/15/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 23:50 Initial Calibration Time(s): 16:30 23:46

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.16	5.16	5.06	5.26	0.00
Aroclor-1248-2 (2)	5.40	5.40	5.30	5.50	0.00
Aroclor-1248-3 (3)	5.44	5.44	5.34	5.54	0.00
Aroclor-1248-4 (4)	5.62	5.62	5.52	5.72	0.00
Aroclor-1248-5 (5)	6.01	6.02	5.92	6.12	0.01
Tetrachloro-m-xylene	4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl	9.21	9.22	9.12	9.32	0.01