

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 10/14/2024

Lab Sample No.: AR1660CCC500 Data File : PP067774.D Time Analyzed: 09:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.922	5.819	6.019	511.340	500.000	2.3
Aroclor-1016-2	5.944	5.842	6.042	500.800	500.000	0.2
Aroclor-1016-3	6.008	5.905	6.105	508.780	500.000	1.8
Aroclor-1016-4	6.106	6.003	6.203	496.670	500.000	-0.7
Aroclor-1016-5	6.400	6.298	6.498	499.480	500.000	-0.1
Aroclor-1260-1	7.525	7.422	7.622	473.850	500.000	-5.2
Aroclor-1260-2	7.778	7.675	7.875	484.720	500.000	-3.1
Aroclor-1260-3	8.140	8.037	8.237	468.630	500.000	-6.3
Aroclor-1260-4	8.378	8.275	8.475	474.560	500.000	-5.1
Aroclor-1260-5	8.716	8.612	8.812	488.110	500.000	-2.4
Decachlorobiphenyl	10.677	10.569	10.769	47.100	50.000	-5.8
Tetrachloro-m-xylene	4.757	4.655	4.855	51.370	50.000	2.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL01 Date Analyzed: 10/14/2024

Lab Sample No.: AR1660CCC500 Data File : PP067774.D Time Analyzed: 09:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.161	5.061	5.261	491.340	500.000	-1.7
Aroclor-1016-2	5.181	5.081	5.281	491.950	500.000	-1.6
Aroclor-1016-3	5.362	5.262	5.462	501.590	500.000	0.3
Aroclor-1016-4	5.401	5.301	5.501	493.990	500.000	-1.2
Aroclor-1016-5	5.621	5.520	5.720	489.800	500.000	-2.0
Aroclor-1260-1	6.664	6.564	6.764	467.300	500.000	-6.5
Aroclor-1260-2	6.851	6.750	6.950	468.690	500.000	-6.3
Aroclor-1260-3	7.009	6.908	7.108	470.960	500.000	-5.8
Aroclor-1260-4	7.484	7.383	7.583	462.780	500.000	-7.4
Aroclor-1260-5	7.722	7.622	7.822	460.480	500.000	-7.9
Decachlorobiphenyl	9.220	9.119	9.319	46.320	50.000	-7.4
Tetrachloro-m-xylene	4.052	3.952	4.152	50.190	50.000	0.4