

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 10/14/2024

Lab Sample No.: AR1660CCC500 Data File : PP067791.D Time Analyzed: 15:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.921	5.819	6.019	532.300	500.000	6.5
Aroclor-1016-2	5.944	5.842	6.042	494.450	500.000	-1.1
Aroclor-1016-3	6.007	5.905	6.105	488.800	500.000	-2.2
Aroclor-1016-4	6.105	6.003	6.203	474.520	500.000	-5.1
Aroclor-1016-5	6.400	6.298	6.498	486.600	500.000	-2.7
Aroclor-1260-1	7.525	7.422	7.622	462.200	500.000	-7.6
Aroclor-1260-2	7.777	7.675	7.875	466.570	500.000	-6.7
Aroclor-1260-3	8.139	8.037	8.237	471.690	500.000	-5.7
Aroclor-1260-4	8.378	8.275	8.475	473.740	500.000	-5.3
Aroclor-1260-5	8.714	8.612	8.812	481.620	500.000	-3.7
Decachlorobiphenyl	10.673	10.569	10.769	49.040	50.000	-1.9
Tetrachloro-m-xylene	4.757	4.655	4.855	51.490	50.000	3.0

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL02 Date Analyzed: 10/14/2024

Lab Sample No.: AR1660CCC500 Data File : PP067791.D Time Analyzed: 15:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.162	5.061	5.261	475.030	500.000	-5.0
Aroclor-1016-2	5.182	5.081	5.281	474.360	500.000	-5.1
Aroclor-1016-3	5.362	5.262	5.462	480.490	500.000	-3.9
Aroclor-1016-4	5.402	5.301	5.501	476.100	500.000	-4.8
Aroclor-1016-5	5.621	5.520	5.720	491.940	500.000	-1.6
Aroclor-1260-1	6.666	6.564	6.764	446.780	500.000	-10.6
Aroclor-1260-2	6.852	6.750	6.950	459.610	500.000	-8.1
Aroclor-1260-3	7.009	6.908	7.108	462.740	500.000	-7.5
Aroclor-1260-4	7.485	7.383	7.583	462.720	500.000	-7.5
Aroclor-1260-5	7.723	7.622	7.822	462.990	500.000	-7.4
Decachlorobiphenyl	9.219	9.119	9.319	46.670	50.000	-6.7
Tetrachloro-m-xylene	4.053	3.952	4.152	48.730	50.000	-2.5