

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL03 Date Analyzed: 10/14/2024

Lab Sample No.: AR1660CCC500 Data File : PP067813.D Time Analyzed: 22:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.917	5.819	6.019	515.980	500.000	3.2
Aroclor-1016-2	5.941	5.842	6.042	495.940	500.000	-0.8
Aroclor-1016-3	6.003	5.905	6.105	502.410	500.000	0.5
Aroclor-1016-4	6.102	6.003	6.203	490.930	500.000	-1.8
Aroclor-1016-5	6.396	6.298	6.498	491.780	500.000	-1.6
Aroclor-1260-1	7.520	7.422	7.622	467.330	500.000	-6.5
Aroclor-1260-2	7.773	7.675	7.875	471.760	500.000	-5.6
Aroclor-1260-3	8.135	8.037	8.237	472.990	500.000	-5.4
Aroclor-1260-4	8.373	8.275	8.475	466.590	500.000	-6.7
Aroclor-1260-5	8.710	8.612	8.812	473.370	500.000	-5.3
Decachlorobiphenyl	10.667	10.569	10.769	48.170	50.000	-3.7
Tetrachloro-m-xylene	4.754	4.655	4.855	50.940	50.000	1.9

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL03 Date Analyzed: 10/14/2024

Lab Sample No.: AR1660CCC500 Data File : PP067813.D Time Analyzed: 22:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.160	5.061	5.261	479.770	500.000	-4.0
Aroclor-1016-2	5.179	5.081	5.281	477.630	500.000	-4.5
Aroclor-1016-3	5.360	5.262	5.462	483.960	500.000	-3.2
Aroclor-1016-4	5.399	5.301	5.501	478.030	500.000	-4.4
Aroclor-1016-5	5.618	5.520	5.720	495.750	500.000	-0.9
Aroclor-1260-1	6.662	6.564	6.764	468.780	500.000	-6.2
Aroclor-1260-2	6.849	6.750	6.950	472.570	500.000	-5.5
Aroclor-1260-3	7.006	6.908	7.108	467.440	500.000	-6.5
Aroclor-1260-4	7.481	7.383	7.583	461.970	500.000	-7.6
Aroclor-1260-5	7.719	7.622	7.822	461.020	500.000	-7.8
Decachlorobiphenyl	9.215	9.119	9.319	46.350	50.000	-7.3
Tetrachloro-m-xylene	4.051	3.952	4.152	48.370	50.000	-3.3