

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL04 Date Analyzed: 10/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP067825.D Time Analyzed: 09:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.918	5.819	6.019	504.470	500.000	0.9
Aroclor-1016-2	5.939	5.842	6.042	495.590	500.000	-0.9
Aroclor-1016-3	6.003	5.905	6.105	485.130	500.000	-3.0
Aroclor-1016-4	6.101	6.003	6.203	482.020	500.000	-3.6
Aroclor-1016-5	6.396	6.298	6.498	498.860	500.000	-0.2
Aroclor-1260-1	7.519	7.422	7.622	510.610	500.000	2.1
Aroclor-1260-2	7.773	7.675	7.875	497.120	500.000	-0.6
Aroclor-1260-3	8.134	8.037	8.237	496.870	500.000	-0.6
Aroclor-1260-4	8.372	8.275	8.475	488.420	500.000	-2.3
Aroclor-1260-5	8.709	8.612	8.812	491.890	500.000	-1.6
Decachlorobiphenyl	10.666	10.569	10.769	49.880	50.000	-0.2
Tetrachloro-m-xylene	4.754	4.655	4.855	51.640	50.000	3.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL04 Date Analyzed: 10/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP067825.D Time Analyzed: 09:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.160	5.061	5.261	484.180	500.000	-3.2
Aroclor-1016-2	5.180	5.081	5.281	485.730	500.000	-2.9
Aroclor-1016-3	5.360	5.262	5.462	495.670	500.000	-0.9
Aroclor-1016-4	5.399	5.301	5.501	489.210	500.000	-2.2
Aroclor-1016-5	5.618	5.520	5.720	537.700	500.000	7.5
Aroclor-1260-1	6.661	6.564	6.764	485.280	500.000	-2.9
Aroclor-1260-2	6.848	6.750	6.950	482.190	500.000	-3.6
Aroclor-1260-3	7.006	6.908	7.108	499.050	500.000	-0.2
Aroclor-1260-4	7.480	7.383	7.583	488.900	500.000	-2.2
Aroclor-1260-5	7.718	7.622	7.822	486.860	500.000	-2.6
Decachlorobiphenyl	9.214	9.119	9.319	48.300	50.000	-3.4
Tetrachloro-m-xylene	4.051	3.952	4.152	49.260	50.000	-1.5