

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL05 Date Analyzed: 10/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP067840.D Time Analyzed: 13:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.917	5.819	6.019	517.170	500.000	3.4
Aroclor-1016-2	5.940	5.842	6.042	502.110	500.000	0.4
Aroclor-1016-3	6.003	5.905	6.105	503.950	500.000	0.8
Aroclor-1016-4	6.102	6.003	6.203	495.350	500.000	-0.9
Aroclor-1016-5	6.396	6.298	6.498	507.580	500.000	1.5
Aroclor-1260-1	7.520	7.422	7.622	502.050	500.000	0.4
Aroclor-1260-2	7.773	7.675	7.875	498.100	500.000	-0.4
Aroclor-1260-3	8.135	8.037	8.237	489.130	500.000	-2.2
Aroclor-1260-4	8.373	8.275	8.475	478.190	500.000	-4.4
Aroclor-1260-5	8.710	8.612	8.812	418.960	500.000	-16.2
Decachlorobiphenyl	10.668	10.569	10.769	49.470	50.000	-1.1
Tetrachloro-m-xylene	4.753	4.655	4.855	51.290	50.000	2.6

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL05 Date Analyzed: 10/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP067840.D Time Analyzed: 13:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.158	5.061	5.261	471.080	500.000	-5.8
Aroclor-1016-2	5.178	5.081	5.281	470.860	500.000	-5.8
Aroclor-1016-3	5.359	5.262	5.462	468.210	500.000	-6.4
Aroclor-1016-4	5.399	5.301	5.501	460.600	500.000	-7.9
Aroclor-1016-5	5.617	5.520	5.720	469.680	500.000	-6.1
Aroclor-1260-1	6.662	6.564	6.764	477.890	500.000	-4.4
Aroclor-1260-2	6.847	6.750	6.950	473.860	500.000	-5.2
Aroclor-1260-3	7.006	6.908	7.108	474.270	500.000	-5.1
Aroclor-1260-4	7.480	7.383	7.583	462.000	500.000	-7.6
Aroclor-1260-5	7.719	7.622	7.822	445.900	500.000	-10.8
Decachlorobiphenyl	9.214	9.119	9.319	47.090	50.000	-5.8
Tetrachloro-m-xylene	4.051	3.952	4.152	47.860	50.000	-4.3