

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL06 Date Analyzed: 10/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP067865.D Time Analyzed: 20:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.917	5.819	6.019	531.810	500.000	6.4
Aroclor-1016-2	5.944	5.842	6.042	576.610	500.000	15.3
Aroclor-1016-3	6.003	5.905	6.105	520.950	500.000	4.2
Aroclor-1016-4	6.101	6.003	6.203	520.890	500.000	4.2
Aroclor-1016-5	6.396	6.298	6.498	524.100	500.000	4.8
Aroclor-1260-1	7.520	7.422	7.622	489.810	500.000	-2.0
Aroclor-1260-2	7.772	7.675	7.875	485.520	500.000	-2.9
Aroclor-1260-3	8.134	8.037	8.237	485.080	500.000	-3.0
Aroclor-1260-4	8.372	8.275	8.475	484.150	500.000	-3.2
Aroclor-1260-5	8.710	8.612	8.812	486.460	500.000	-2.7
Decachlorobiphenyl	10.666	10.569	10.769	49.890	50.000	-0.2
Tetrachloro-m-xylene	4.753	4.655	4.855	52.150	50.000	4.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL06 Date Analyzed: 10/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP067865.D Time Analyzed: 20:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.159	5.061	5.261	521.720	500.000	4.3
Aroclor-1016-2	5.179	5.081	5.281	506.320	500.000	1.3
Aroclor-1016-3	5.359	5.262	5.462	547.220	500.000	9.4
Aroclor-1016-4	5.398	5.301	5.501	530.880	500.000	6.2
Aroclor-1016-5	5.618	5.520	5.720	554.590	500.000	10.9
Aroclor-1260-1	6.661	6.564	6.764	479.920	500.000	-4.0
Aroclor-1260-2	6.847	6.750	6.950	477.330	500.000	-4.5
Aroclor-1260-3	7.006	6.908	7.108	475.190	500.000	-5.0
Aroclor-1260-4	7.480	7.383	7.583	467.600	500.000	-6.5
Aroclor-1260-5	7.718	7.622	7.822	468.400	500.000	-6.3
Decachlorobiphenyl	9.213	9.119	9.319	46.520	50.000	-7.0
Tetrachloro-m-xylene	4.051	3.952	4.152	50.900	50.000	1.8