



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4397

Client: Portal Partners Tri-Venture

Analytical Method: 8082A Datafile : PP067788.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB164124BS	AR1016	166.5	153	ug/kg	92				40 (71)	140 (120)	
	AR1260	166.5	143	ug/kg	86				40 (65)	140 (130)	