



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4397

Client: Portal Partners Tri-Venture

Analytical Method: **8082A** **Datafile :** PP067846.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB164139BSD	AR1016	5	4.30	ug/L	86	7			40 (61)	140 (112)	20 (20)
	AR1260	5	4.10	ug/L	82	5			40 (66)	140 (113)	20 (20)