



Contract:	<u>PORT06</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>P4397</u>	SAS No.:	<u>P4397</u>	SDG NO.:	<u>P4397</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):		<u>10/23/2024</u>	<u>10/23/2024</u>		
		Calibration Times:		<u>11:28</u>	<u>13:04</u>		

LAB FILE ID:		RT 200 =	<u>PS028008.D</u>	RT 500 =	<u>PS028009.D</u>
RT 750 =	<u>PS028010.D</u>	RT 1000 =	<u>PS028011.D</u>	RT 1500 =	<u>PS028012.D</u>

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

Instrument ID: ECD_S Calibration Date(s): 10/23/2024 10/23/2024

Calibration Times: 11:28 13:04

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS028008.D</u>	RT 500 = <u>PS028009.D</u>
	RT 750 = <u>PS028010.D</u>	RT 1000 = <u>PS028011.D</u>
		RT 1500 = <u>PS028012.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM TO	
2,4,5-TP(Silvex)	9.73	9.73	9.73	9.73	9.72	9.73	9.63	9.83
2,4-D	8.84	8.84	8.84	8.84	8.84	8.84	8.74	8.94
2,4-DCAA	7.61	7.62	7.61	7.62	7.61	7.61	7.51	7.71