

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM **Case No.:** P4397 **SAS No.:** P4397 **SDG NO.:** P4397

Continuing Calib Date: 10/24/2024 **Initial Calibration Date(s):** 10/23/2024 10/23/2024

Continuing Calib Time: 09:50 **Initial Calibration Time(s):** 11:28 13:04

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.09	7.09	6.99	7.19	0.00
2,4-D	8.18	8.18	8.08	8.28	0.00
2,4,5-TP(Silvex)	9.03	9.03	8.93	9.13	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM **Case No.:** P4397 **SAS No.:** P4397 **SDG NO.:** P4397

Continuing Calib Date: 10/24/2024 **Initial Calibration Date(s):** 10/23/2024 10/23/2024

Continuing Calib Time: 09:50 **Initial Calibration Time(s):** 11:28 13:04

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.61	7.61	7.51	7.71	0.00
2,4-D	8.84	8.84	8.74	8.94	0.00
2,4,5-TP(Silvex)	9.73	9.73	9.63	9.83	0.00