

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL02 Date Analyzed: 10/24/2024

Lab Sample No.: HSTDCCC750 Data File : PS028047.D Time Analyzed: 21:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.027	8.929	9.129	693.140	712.500	-2.7
2,4-D	8.176	8.077	8.277	681.790	705.000	-3.3
2,4-DCAA	7.090	6.991	7.191	722.190	750.000	-3.7

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/23/2024 10/23/2024

Client Sample No.: CCAL02 Date Analyzed: 10/24/2024

Lab Sample No.: HSTDCCC750 Data File : PS028047.D Time Analyzed: 21:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.723	9.625	9.825	791.420	712.500	11.1
2,4-D	8.833	8.737	8.937	737.820	705.000	4.7
2,4-DCAA	7.612	7.514	7.714	784.260	750.000	4.6