

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102424\
 Data File : PS028036.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Oct 2024 11:25
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 01:11:58 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102324.M
 Quant Title : 8080.M
 QLast Update : Wed Oct 23 13:25:49 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.091	7.613	1914.7E6	695.5E6	709.955	733.178
Target Compounds							
1) T	Dalapon	2.541	2.609	2926.4E6	1183.0E6	642.263	693.684
2) T	3,5-DICHL...	6.283	6.586	2604.7E6	927.7E6	658.246	693.571
3) T	4-Nitroph...	6.885	7.147	1187.3E6	453.5E6	641.664	655.989
5) T	DICAMBA	7.269	7.806	7451.1E6	2400.1E6	681.281	726.281
6) T	MCPD	7.449	7.909	510.0E6	177.2E6	69.554	68.553
7) T	MCPA	7.594	8.151	706.2E6	308.2E6	66.602	66.971
8) T	DICHLORPROP	7.955	8.510	1920.5E6	672.1E6	665.995	717.319
9) T	2,4-D	8.176	8.835	2255.3E6	707.9E6	665.579	709.891
10) T	Pentachlo...	8.463	9.347	27226.3E6	9287.8E6	688.019	725.260
11) T	2,4,5-TP ...	9.028	9.725	11061.0E6	3363.3E6	684.860	736.810
12) T	2,4,5-T	9.313	10.139	11345.2E6	3060.3E6	679.090	744.163
13) T	2,4-DB	9.875	10.699	1789.3E6	376.1E6	679.603	736.376
14) T	DINOSEB	11.051	11.073	7315.6E6	2351.4E6	664.638	698.821
15) T	Picloram	10.866	12.146	14972.7E6	3313.7E6	672.468	727.207
16) T	DCPA	11.350	12.104	12750.2E6	3536.9E6	689.629	762.079

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS102424\
Data File : PS028036.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Oct 2024 11:25
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 01:11:58 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS102324.M
Quant Title : 8080.M
QLast Update : Wed Oct 23 13:25:49 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

