

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2024 10/21/2024

Client Sample No.: CCAL04 Date Analyzed: 10/25/2024

Lab Sample No.: PSTDCCC050 Data File : PL092650.D Time Analyzed: 19:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.058	8.955	9.155	49.750	50.000	-0.5
Endrin	6.577	6.474	6.674	41.530	50.000	-16.9
gamma-BHC (Lindane)	4.330	4.227	4.427	46.750	50.000	-6.5
Heptachlor	4.918	4.815	5.015	45.090	50.000	-9.8
Heptachlor epoxide	5.686	5.583	5.783	43.960	50.000	-12.1
Methoxychlor	7.501	7.400	7.600	49.670	50.000	-0.7
Tetrachloro-m-xylene	3.541	3.438	3.638	49.590	50.000	-0.8

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/21/2024 10/21/2024

Client Sample No.: CCAL04 Date Analyzed: 10/25/2024

Lab Sample No.: PSTDCCC050 Data File : PL092650.D Time Analyzed: 19:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.918	7.817	8.017	54.870	50.000	9.7
Endrin	5.644	5.541	5.741	51.490	50.000	3.0
gamma-BHC (Lindane)	3.612	3.510	3.710	54.770	50.000	9.5
Heptachlor	3.951	3.849	4.049	53.610	50.000	7.2
Heptachlor epoxide	4.734	4.631	4.831	54.180	50.000	8.4
Methoxychlor	6.617	6.515	6.715	53.570	50.000	7.1
Tetrachloro-m-xylene	2.779	2.677	2.877	55.300	50.000	10.6