

Surrogate Summary

SDG No.: **P4397**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL092494.D	PIBLK-PL092494.D	Decachlorobiphenyl	1	20	21.5	107		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	22.0	110		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.4	107		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.0	100		30 (77)	150 (126)
I.BLK-PL092589.D	PIBLK-PL092589.D	Decachlorobiphenyl	1	20	21.0	105		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.0	100		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	23.3	116		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.0	105		30 (77)	150 (126)
PB164360BL	PB164360BL	Decachlorobiphenyl	1	20	19.8	99		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.6	93		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.0	105		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.5	98		30 (77)	150 (126)
PB164261TB	PB164261TB	Decachlorobiphenyl	1	20	19.9	100		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.3	91		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.6	108		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.1	96		30 (77)	150 (126)
P4397-06	WB-301-BOT	Decachlorobiphenyl	1	20	20.6	103		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	17.5	88		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.6	108		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	16.6	83		30 (77)	150 (126)
P4397-06MS	WB-301-BOTMS	Decachlorobiphenyl	1	20	19.4	97		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	15.9	80		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	20.9	105		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	16.3	82		30 (77)	150 (126)
P4397-06MSD	WB-301-BOTMSD	Decachlorobiphenyl	1	20	19.9	100		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	16.5	83		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.6	108		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	16.5	83		30 (77)	150 (126)
I.BLK-PL092612.D	PIBLK-PL092612.D	Decachlorobiphenyl	1	20	22.3	111		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	21.4	107		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	23.4	117		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	22.1	111		30 (77)	150 (126)
I.BLK-PL092637.D	PIBLK-PL092637.D	Decachlorobiphenyl	1	20	22.1	110		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.9	100		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	24.4	122		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.6	108		30 (77)	150 (126)
PB164360BS	PB164360BS	Decachlorobiphenyl	1	20	17.7	88		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	15.7	79		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	19.5	97		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	16.3	81		30 (77)	150 (126)
I.BLK-PL092649.D	PIBLK-PL092649.D	Decachlorobiphenyl	1	20	21.2	106		30 (43)	150 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: P4397

Client: Portal Partners Tri-Venture

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL092649.D	PIBLK-PL092649.D	Tetrachloro-m-xylene	1	20	20.0	100		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	22.2	111		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.4	107		30 (77)	150 (126)