

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG No.: P4397

Instrument ID: BNA_F Calibration Date/Time: 10/22/2024 14:28

Lab File ID: BF139927.D Init. Calib. Date(s): 10/18/2024 10/18/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:27 13:46

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.278	1.238		-3.1	
Benzaldehyde	0.931	0.957		2.8	
Phenol-d6	1.655	1.570		-5.1	
Phenol	1.706	1.639		-3.9	20.0
bis(2-Chloroethyl)ether	1.319	1.272		-3.6	
2-Chlorophenol	1.306	1.287		-1.5	
2-Methylphenol	1.114	1.070		-4.0	
2,2-oxybis(1-Chloropropane)	2.248	2.090		-7.0	
Acetophenone	0.490	0.474		-3.3	
3+4-Methylphenols	1.424	1.367		-4.0	
n-Nitroso-di-n-propylamine	1.004	0.934	0.050	-7.0	
Nitrobenzene-d5	0.361	0.377		4.4	
Hexachloroethane	0.534	0.531		-0.6	
Nitrobenzene	0.396	0.401		1.3	
Isophorone	0.685	0.658		-3.9	
2-Nitrophenol	0.149	0.178		19.5	20.0
2,4-Dimethylphenol	0.249	0.243		-2.4	
bis(2-Chloroethoxy)methane	0.415	0.406		-2.2	
2,4-Dichlorophenol	0.283	0.285		0.7	20.0
Naphthalene	1.031	1.017		-1.4	
4-Chloroaniline	0.352	0.346		-1.7	
Hexachlorobutadiene	0.197	0.204		3.6	20.0
Caprolactam	0.090	0.086		-4.4	
4-Chloro-3-methylphenol	0.314	0.306		-2.5	20.0
2-Methylnaphthalene	0.632	0.625		-1.1	
Hexachlorocyclopentadiene	0.188	0.217	0.050	15.4	
2,4,6-Trichlorophenol	0.382	0.396		3.7	20.0
2-Fluorobiphenyl	1.210	1.202		-0.7	
2,4,5-Trichlorophenol	0.394	0.423		7.4	
1,1-Biphenyl	1.392	1.411		1.4	
2-Chloronaphthalene	1.121	1.140		1.7	
2-Nitroaniline	0.343	0.368		7.3	
Dimethylphthalate	1.246	1.230		-1.3	
Acenaphthylene	1.621	1.627		0.4	
2,6-Dinitrotoluene	0.270	0.289		7.0	
3-Nitroaniline	0.278	0.287		3.2	
Acenaphthene	1.049	1.066		1.6	20.0
2,4-Dinitrophenol	0.093	0.132	0.050	41.9	
4-Nitrophenol	0.214	0.223	0.050	4.2	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG No.: P4397

Instrument ID: BNA_F Calibration Date/Time: 10/22/2024 14:28

Lab File ID: BF139927.D Init. Calib. Date(s): 10/18/2024 10/18/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:27 13:46

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.505	1.499		-0.4	
2,4-Dinitrotoluene	0.332	0.373		12.3	
Diethylphthalate	1.223	1.185		-3.1	
4-Chlorophenyl-phenylether	0.579	0.580		0.2	
Fluorene	1.146	1.117		-2.5	
4-Nitroaniline	0.263	0.268		1.9	
4,6-Dinitro-2-methylphenol	0.082	0.112		36.6	
n-Nitrosodiphenylamine	0.599	0.598		-0.2	20.0
2,4,6-Tribromophenol	0.187	0.201		7.5	
4-Bromophenyl-phenylether	0.206	0.216		4.9	
Hexachlorobenzene	0.232	0.239		3.0	
Atrazine	0.162	0.155		-4.3	
Pentachlorophenol	0.141	0.160		13.5	20.0
Phenanthrene	0.945	0.925		-2.1	
Anthracene	0.922	0.916		-0.7	
Carbazole	0.857	0.810		-5.5	
Di-n-butylphthalate	0.990	0.919		-7.2	
Fluoranthene	0.955	0.875		-8.5	20.0
Pyrene	1.751	1.852		5.8	
Terphenyl-d14	1.227	1.269		3.4	
Butylbenzylphthalate	0.525	0.539		2.7	
3,3-Dichlorobenzidine	0.380	0.452		18.9	
Benzo (a) anthracene	1.302	1.303		0.1	
Chrysene	1.194	1.167		-2.3	
Bis(2-ethylhexyl)phthalate	0.589	0.693		17.7	
Di-n-octyl phthalate	1.071	1.312		22.5	20.0
Benzo (b) fluoranthene	1.218	1.229		0.9	
Benzo (k) fluoranthene	1.051	0.899		-14.5	
Benzo (a) pyrene	1.002	1.008		0.6	20.0
Indeno (1,2,3-cd) pyrene	1.287	1.437		11.7	
Dibenzo (a,h) anthracene	1.073	1.186		10.5	
Benzo (g,h,i) perylene	1.072	1.213		13.2	
1,2,4,5-Tetrachlorobenzene	0.540	0.557		3.1	
1,4-Dioxane	0.596	0.561		-5.9	20.0
2,3,4,6-Tetrachlorophenol	0.309	0.321		3.9	

All other compounds must meet a minimum RRF of 0.010.