

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG No.: P4397

Instrument ID: BNA_F Calibration Date/Time: 10/23/2024 15:30

Lab File ID: BF139965.D Init. Calib. Date(s): 10/18/2024 10/18/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:27 13:46

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.278	1.233		-3.5	
Benzaldehyde	0.931	0.939		0.9	
Phenol-d6	1.655	1.579		-4.6	
Phenol	1.706	1.663		-2.5	20.0
bis(2-Chloroethyl)ether	1.319	1.273		-3.5	
2-Chlorophenol	1.306	1.291		-1.1	
2-Methylphenol	1.114	1.099		-1.3	
2,2-oxybis(1-Chloropropane)	2.248	2.018		-10.2	
Acetophenone	0.490	0.467		-4.7	
3+4-Methylphenols	1.424	1.388		-2.5	
n-Nitroso-di-n-propylamine	1.004	0.918	0.050	-8.6	
Nitrobenzene-d5	0.361	0.370		2.5	
Hexachloroethane	0.534	0.530		-0.7	
Nitrobenzene	0.396	0.391		-1.3	
Isophorone	0.685	0.651		-5.0	
2-Nitrophenol	0.149	0.175		17.5	20.0
2,4-Dimethylphenol	0.249	0.239		-4.0	
bis(2-Chloroethoxy)methane	0.415	0.397		-4.3	
2,4-Dichlorophenol	0.283	0.283		0.0	20.0
Naphthalene	1.031	1.011		-1.9	
4-Chloroaniline	0.352	0.340		-3.4	
Hexachlorobutadiene	0.197	0.199		1.0	20.0
Caprolactam	0.090	0.092		2.2	
4-Chloro-3-methylphenol	0.314	0.308		-1.9	20.0
2-Methylnaphthalene	0.632	0.618		-2.2	
Hexachlorocyclopentadiene	0.188	0.193	0.050	2.7	
2,4,6-Trichlorophenol	0.382	0.381		-0.3	20.0
2-Fluorobiphenyl	1.210	1.169		-3.4	
2,4,5-Trichlorophenol	0.394	0.415		5.3	
1,1-Biphenyl	1.392	1.351		-2.9	
2-Chloronaphthalene	1.121	1.097		-2.1	
2-Nitroaniline	0.343	0.369		7.6	
Dimethylphthalate	1.246	1.226		-1.6	
Acenaphthylene	1.621	1.588		-2.0	
2,6-Dinitrotoluene	0.270	0.289		7.0	
3-Nitroaniline	0.278	0.292		5.0	
Acenaphthene	1.049	1.051		0.2	20.0
2,4-Dinitrophenol	0.093	0.139	0.050	49.5	
4-Nitrophenol	0.214	0.228	0.050	6.5	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG No.: P4397

Instrument ID: BNA_F Calibration Date/Time: 10/23/2024 15:30

Lab File ID: BF139965.D Init. Calib. Date(s): 10/18/2024 10/18/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:27 13:46

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.505	1.458		-3.1	
2,4-Dinitrotoluene	0.332	0.379		14.2	
Diethylphthalate	1.223	1.211		-1.0	
4-Chlorophenyl-phenylether	0.579	0.574		-0.9	
Fluorene	1.146	1.118		-2.4	
4-Nitroaniline	0.263	0.280		6.5	
4,6-Dinitro-2-methylphenol	0.082	0.112		36.6	
n-Nitrosodiphenylamine	0.599	0.583		-2.7	20.0
2,4,6-Tribromophenol	0.187	0.204		9.1	
4-Bromophenyl-phenylether	0.206	0.211		2.4	
Hexachlorobenzene	0.232	0.234		0.9	
Atrazine	0.162	0.166		2.5	
Pentachlorophenol	0.141	0.157		11.3	20.0
Phenanthrene	0.945	0.920		-2.6	
Anthracene	0.922	0.894		-3.0	
Carbazole	0.857	0.815		-4.9	
Di-n-butylphthalate	0.990	0.956		-3.4	
Fluoranthene	0.955	0.893		-6.5	20.0
Pyrene	1.751	1.731		-1.1	
Terphenyl-d14	1.227	1.227		0.0	
Butylbenzylphthalate	0.525	0.540		2.9	
3,3-Dichlorobenzidine	0.380	0.447		17.6	
Benzo (a) anthracene	1.302	1.293		-0.7	
Chrysene	1.194	1.157		-3.1	
Bis(2-ethylhexyl)phthalate	0.589	0.666		13.1	
Di-n-octyl phthalate	1.071	1.198		11.9	20.0
Benzo (b) fluoranthene	1.218	1.083		-11.1	
Benzo (k) fluoranthene	1.051	1.048		-0.3	
Benzo (a) pyrene	1.002	0.978		-2.4	20.0
Indeno (1,2,3-cd) pyrene	1.287	1.280		-0.5	
Dibenzo (a,h) anthracene	1.073	1.060		-1.2	
Benzo (g,h,i) perylene	1.072	1.053		-1.8	
1,2,4,5-Tetrachlorobenzene	0.540	0.540		0.0	
1,4-Dioxane	0.596	0.581		-2.5	20.0
2,3,4,6-Tetrachlorophenol	0.309	0.324		4.9	

All other compounds must meet a minimum RRF of 0.010.