

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bf102124

SequenceID :	<u>bf102124</u>	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)		_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)		_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series		_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series		_____	_____	_____✓_____
5. GC/MS Calibration Met:		_____	_____	_____✓_____
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.		_____	_____	_____✓_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:		_____	_____✓_____	_____
a. B/N Fraction				
d. Acid Fraction				

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

_____ ✓ _____

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

_____ ✓ _____

a. B/N Fraction Recovery and RPD failed for some compound in P4385-06MS/MSD and P4419-01MS/MSD due to matrix interference. Recovery and RPD failed for certain compounds even in the diluted sample of P4410-01MS/MSD due to matrix interference. Therefore no corrective action is required.

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

_____ ✓ _____

Internal Standard Failed in the sample P4430-01 which is not associated for required compound.

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

_____ _____ ✓

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

_____ _____ ✓

ADDITIONAL COMMENTS:

The samples P4410-01 and its MS/MSD, P4400-01, P4406-01, P4455-01 had to be analyzed with their respective dilutions due to dirty and viscous matrix which was difficult to inject otherwise. Hence no corrective action is required.

Analyst

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: bf102124

SequenceID : bf102124 NA NO YES