

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/21/2024

Lab File ID: BF139892.D Time Analyzed: 09:57

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	170135	6.892	645389	8.18	355580	9.93
UPPER LIMIT	340270	7.392	1290780	8.675	711160	10.427
LOWER LIMIT	85067.5	6.392	322695	7.675	177790	9.427
EPA SAMPLE NO.						
01 PB164123BL	167950	6.89	682521	8.17	387913	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.