

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/22/2024

Lab File ID: BF139927.D Time Analyzed: 14:28

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	169434	6.893	628861	8.18	340877	9.93
UPPER LIMIT	338868	7.393	1257720	8.675	681754	10.428
LOWER LIMIT	84717	6.393	314431	7.675	170439	9.428
EPA SAMPLE NO.						
01 PB164154BL	163319	6.89	633200	8.17	357885	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.