

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/21/2024

Lab File ID: BF139892.D Time Analyzed: 09:57

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	607097	11.416	292847	14.057	323975	15.533
UPPER LIMIT	1214190	11.916	585694	14.557	647950	16.033
LOWER LIMIT	303549	10.916	146424	13.557	161988	15.033
EPA SAMPLE NO.						
01 PB164123BL	717487	11.41	434664	14.05	327943	15.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.