

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102424\
 Data File : BF140001.D
 Acq On : 24 Oct 2024 15:16
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 24 16:03:07 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2	1,4-Dioxane	0.596	0.573	3.9	87	-0.02
3	Pyridine	1.476	1.408	4.6	84	-0.01
4	n-Nitrosodimethylamine	0.793	0.688	13.2	76	-0.01
5 S	2-Fluorophenol	1.278	1.208	5.5	84	0.00
6	Aniline	1.542	1.435	6.9	81	0.00
7 S	Phenol-d6	1.655	1.521	8.1	83	0.00
8	2-Chlorophenol	1.306	1.259	3.6	86	0.00
9	Benzaldehyde	0.931	0.931	0.0	88	0.00
10 C	Phenol	1.706	1.601	6.2	84	0.00
11	bis(2-Chloroethyl)ether	1.319	1.259	4.5	87	0.00
12	1,3-Dichlorobenzene	1.499	1.462	2.5	87	0.00
13 C	1,4-Dichlorobenzene	1.498	1.466	2.1	88	0.00
14	1,2-Dichlorobenzene	1.398	1.368	2.1	89	0.00
15	Benzyl Alcohol	1.214	1.131	6.8	83	0.00
16	2,2'-oxybis(1-Chloropropane	2.248	1.948	13.3	77	0.00
17	2-Methylphenol	1.114	1.047	6.0	84	0.00
18	Hexachloroethane	0.534	0.526	1.5	87	0.00
19 P	n-Nitroso-di-n-propylamine	1.004	0.889	11.5	81	0.00
20	3+4-Methylphenols	1.424	1.322	7.2	83	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
22	Acetophenone	0.490	0.472	3.7	84	0.00
23 S	Nitrobenzene-d5	0.361	0.374	-3.6	87	0.00
24	Nitrobenzene	0.396	0.397	-0.3	85	0.00
25	Isophorone	0.685	0.655	4.4	83	0.00
26 C	2-Nitrophenol	0.149	0.178	-19.5	98	0.00
27	2,4-Dimethylphenol	0.249	0.238	4.4	83	0.00
28	bis(2-Chloroethoxy)methane	0.415	0.407	1.9	85	0.00
29 C	2,4-Dichlorophenol	0.283	0.285	-0.7	87	0.00
30	1,2,4-Trichlorobenzene	0.314	0.319	-1.6	87	0.00
31	Naphthalene	1.031	1.015	1.6	86	0.00
32	Benzoic acid	0.217	0.241	-11.1	95	0.00
33	4-Chloroaniline	0.352	0.336	4.5	82	0.00
34 C	Hexachlorobutadiene	0.197	0.203	-3.0	89	0.00
35	Caprolactam	0.090	0.089	1.1	83	0.00
36 C	4-Chloro-3-methylphenol	0.314	0.306	2.5	84	0.00
37	2-Methylnaphthalene	0.632	0.623	1.4	86	0.00
38	1-Methylnaphthalene	0.620	0.608	1.9	86	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	0.540	0.560	-3.7	89	0.00
41 P	Hexachlorocyclopentadiene	0.188	0.199	-5.9	86	0.00
42 S	2,4,6-Tribromophenol	0.187	0.207	-10.7	95	0.00
43 C	2,4,6-Trichlorophenol	0.382	0.387	-1.3	86	0.00
44	2,4,5-Trichlorophenol	0.394	0.415	-5.3	89	0.00
45 S	2-Fluorobiphenyl	1.210	1.197	1.1	88	0.00
46	1,1'-Biphenyl	1.392	1.393	-0.1	87	0.00
47	2-Chloronaphthalene	1.121	1.127	-0.5	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102424\
 Data File : BF140001.D
 Acq On : 24 Oct 2024 15:16
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 24 16:03:07 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 18 15:07:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.343	0.369	-7.6	87	0.00
49	Acenaphthylene	1.621	1.600	1.3	85	0.00
50	Dimethylphthalate	1.246	1.232	1.1	85	0.00
51	2,6-Dinitrotoluene	0.270	0.292	-8.1	89	0.00
52 C	Acenaphthene	1.049	1.061	-1.1	87	0.00
53	3-Nitroaniline	0.278	0.285	-2.5	84	0.00
54 P	2,4-Dinitrophenol	0.093	0.142	-52.7#	120	0.00
55	Dibenzofuran	1.505	1.482	1.5	85	0.00
56 P	4-Nitrophenol	0.214	0.216	-0.9	80	0.00
57	2,4-Dinitrotoluene	0.332	0.371	-11.7	89	0.00
58	Fluorene	1.146	1.115	2.7	86	0.00
59	2,3,4,6-Tetrachlorophenol	0.309	0.321	-3.9	89	0.00
60	Diethylphthalate	1.223	1.183	3.3	83	0.00
61	4-Chlorophenyl-phenylether	0.579	0.577	0.3	87	0.00
62	4-Nitroaniline	0.263	0.268	-1.9	83	0.00
63	Azobenzene	1.297	1.207	6.9	79	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
65	4,6-Dinitro-2-methylphenol	0.082	0.115	-40.2#	110	0.00
66 c	n-Nitrosodiphenylamine	0.599	0.597	0.3	83	0.00
67	4-Bromophenyl-phenylether	0.206	0.220	-6.8	90	0.00
68	Hexachlorobenzene	0.232	0.242	-4.3	87	0.00
69	Atrazine	0.162	0.163	-0.6	78	0.00
70 C	Pentachlorophenol	0.141	0.159	-12.8	87	0.00
71	Phenanthrene	0.945	0.922	2.4	82	0.00
72	Anthracene	0.922	0.910	1.3	83	0.00
73	Carbazole	0.857	0.810	5.5	80	0.00
74	Di-n-butylphthalate	0.990	0.945	4.5	77	0.00
75 C	Fluoranthene	0.955	0.896	6.2	79	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
77	Benzydine	0.329	0.542	-64.7#	113	0.00
78	Pyrene	1.751	1.855	-5.9	79	0.00
79 S	Terphenyl-d14	1.227	1.306	-6.4	80	0.00
80	Butylbenzylphthalate	0.525	0.544	-3.6	75	0.00
81	Benzo(a)anthracene	1.302	1.322	-1.5	76	0.00
82	3,3'-Dichlorobenzidine	0.380	0.435	-14.5	86	0.00
83	Chrysene	1.194	1.125	5.8	74	0.00
84	Bis(2-ethylhexyl)phthalate	0.589	0.627	-6.5	77	0.00
85 c	Di-n-octyl phthalate	1.071	1.102	-2.9	74	0.00
86 I	Perylene-d12	1.000	1.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	1.287	1.386	-7.7	87	0.00
88	Benzo(b)fluoranthene	1.218	1.087	10.8	77	0.00
89	Benzo(k)fluoranthene	1.051	1.032	1.8	82	0.00
90 C	Benzo(a)pyrene	1.002	0.984	1.8	81	0.00
91	Dibenzo(a,h)anthracene	1.073	1.154	-7.5	87	0.00
92	Benzo(g,h,i)perylene	1.072	1.167	-8.9	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF102424\
Data File : BF140001.D
Acq On : 24 Oct 2024 15:16
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Oct 24 16:03:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF101824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 18 15:07:50 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0