

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4397

Client: Portal Partners Tri-Venture

Analytical Method: 8270E DataFile: BF139994.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164315BS	Pyridine	50	38.8	ug/L	78				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	45.3	ug/L	91				70 (76)	130 (103)	
	2-Methylphenol	50	46.6	ug/L	93				70 (69)	130 (109)	
	3+4-Methylphenols	50	44.2	ug/L	88				20 (67)	160 (106)	
	Hexachloroethane	50	45.5	ug/L	91				20 (76)	160 (118)	
	Nitrobenzene	50	43.6	ug/L	87				70 (58)	130 (106)	
	Hexachlorobutadiene	50	45.3	ug/L	91				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	50.9	ug/L	102				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	49.7	ug/L	99				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	54.7	ug/L	109				70 (60)	130 (115)	
	Hexachlorobenzene	50	49.6	ug/L	99				70 (73)	130 (106)	
	Pentachlorophenol	100	99.2	ug/L	99				20 (47)	160 (114)	