

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/24/2024

Lab File ID: BF139990.D Time Analyzed: 09:55

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	160242	6.892	587754	8.18	322730	9.93
UPPER LIMIT	320484	7.392	1175510	8.675	645460	10.428
LOWER LIMIT	80121	6.392	293877	7.675	161365	9.428
EPA SAMPLE NO.						
01 PB164315BL	148560	6.89	574491	8.17	327603	9.93
02 PB164315BS	143702	6.89	566601	8.18	302186	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.