

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/24/2024

Lab File ID: BF140001.D Time Analyzed: 15:16

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	551060	11.416	264546	14.051	308388	15.533
UPPER LIMIT	1102120	11.916	529092	14.551	616776	16.033
LOWER LIMIT	275530	10.916	132273	13.551	154194	15.033
EPA SAMPLE NO.						
01 PB164261TB	603559	11.41	330965	14.05	294008	15.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.