

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: P4397

SDG NO.: P4397

Lab File ID: BF139989.D

DFTPP Injection Date: 10/24/2024

Instrument ID: BNA_F

DFTPP Injection Time: 09:26

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.5
68	Less than 2.0% of mass 69	0.6 (1.9) 1
69	Mass 69 relative abundance	29.8
70	Less than 2.0% of mass 69	0.2 (0.7) 1
127	10.0 - 80.0% of mass 198	37.5
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.6
275	10.0 - 60.0% of mass 198	24.9
365	Greater than 1% of mass 198	3.2
441	Present, but less than mass 443	14.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF139990.D	10/24/2024	09:55
PB164315BL	PB164315BL	BF139993.D	10/24/2024	11:20
PB164315BS	PB164315BS	BF139994.D	10/24/2024	11:49