

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397
Lab File ID: VN084450.D Date Analyzed: 10/22/2024
Instrument ID: MSVOA_N Time Analyzed: 12:50
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	171974	8.22	293014	9.10	264421	11.87
UPPER LIMIT	343948	8.724	586028	9.6	528842	12.365
LOWER LIMIT	85987	7.724	146507	8.6	132211	11.365
EPA SAMPLE NO.						
WB-301-BOT	168250	8.22	297819	9.10	273036	11.87
VN1022WBL01	195991	8.22	325943	9.10	276681	11.87
VN1022WBS01	177710	8.22	304170	9.10	270028	11.87
VN1022WBSD01	162648	8.22	284290	9.10	251894	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.