

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397  
 Lab File ID: VN084450.D Date Analyzed: 10/22/2024  
 Instrument ID: MSVOA\_N Time Analyzed: 12:50  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 131065        | 13.788 |  |  |  |  |
| UPPER LIMIT    | 262130        | 14.288 |  |  |  |  |
| LOWER LIMIT    | 65532.5       | 13.288 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| WB-301-BOT     | 103880        | 13.79  |  |  |  |  |
| VN1022WBL01    | 106919        | 13.79  |  |  |  |  |
| VN1022WBS01    | 132421        | 13.79  |  |  |  |  |
| VN1022WBSD01   | 120786        | 13.79  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.