

Surrogate Summary

SDG No.: **P4397**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4397-01	WB-301-TOP	1,2-Dichloroethane-d4	50	63.1	126	70 (50)	130 (163)
		Dibromofluoromethane	50	51.8	104	70 (54)	130 (147)
		Toluene-d8	50	50.7	101	70 (58)	130 (134)
		4-Bromofluorobenzene	50	43.9	88	70 (29)	130 (146)
P4397-02	WB-301-BOT	1,2-Dichloroethane-d4	50	57.1	114	70 (50)	130 (163)
		Dibromofluoromethane	50	49.7	99	70 (54)	130 (147)
		Toluene-d8	50	50.2	100	70 (58)	130 (134)
		4-Bromofluorobenzene	50	44.1	88	70 (29)	130 (146)
VY1014SBL01	VY1014SBL01	1,2-Dichloroethane-d4	50	52.0	104	70 (50)	130 (163)
		Dibromofluoromethane	50	48.9	98	70 (54)	130 (147)
		Toluene-d8	50	50.2	100	70 (58)	130 (134)
		4-Bromofluorobenzene	50	40.3	81	70 (29)	130 (146)
VY1014SBS01	VY1014SBS01	1,2-Dichloroethane-d4	50	55.1	110	70 (50)	130 (163)
		Dibromofluoromethane	50	55.0	110	70 (54)	130 (147)
		Toluene-d8	50	52.7	105	70 (58)	130 (134)
		4-Bromofluorobenzene	50	53.3	107	70 (29)	130 (146)
VY1014SBSD01	VY1014SBSD01	1,2-Dichloroethane-d4	50	54.7	109	70 (50)	130 (163)
		Dibromofluoromethane	50	54.2	108	70 (54)	130 (147)
		Toluene-d8	50	52.3	105	70 (58)	130 (134)
		4-Bromofluorobenzene	50	51.5	103	70 (29)	130 (146)

Surrogate Summary

SDG No.: **P4397**

Client: **Portal Partners Tri-Venture**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4397-04	WB-301-SW	1,2-Dichloroethane-d4	50	50.3	101	70 (74)	130 (125)
		Dibromofluoromethane	50	47.8	96	70 (75)	130 (124)
		Toluene-d8	50	49.7	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.4	95	70 (77)	130 (121)
P4397-05	TB-10102024	1,2-Dichloroethane-d4	50	50.5	101	70 (74)	130 (125)
		Dibromofluoromethane	50	47.3	95	70 (75)	130 (124)
		Toluene-d8	50	48.9	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.5	95	70 (77)	130 (121)
VX1015WBL01	VX1015WBL01	1,2-Dichloroethane-d4	50	50.4	101	70 (74)	130 (125)
		Dibromofluoromethane	50	47.6	95	70 (75)	130 (124)
		Toluene-d8	50	49.6	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.6	97	70 (77)	130 (121)
VX1015WBS01	VX1015WBS01	1,2-Dichloroethane-d4	50	54.2	108	70 (74)	130 (125)
		Dibromofluoromethane	50	52.3	105	70 (75)	130 (124)
		Toluene-d8	50	50.0	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.2	102	70 (77)	130 (121)
VX1015WBSD0	VX1015WBSD01	1,2-Dichloroethane-d4	50	53.6	107	70 (74)	130 (125)
		Dibromofluoromethane	50	52.6	105	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.5	105	70 (77)	130 (121)

