

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397
 Lab File ID: VX043389.D Date Analyzed: 10/15/2024
 Instrument ID: MSVOA_X Time Analyzed: 08:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	108973	5.54	187563	6.76	162899	10.05
UPPER LIMIT	217946	6.044	375126	7.257	325798	10.549
LOWER LIMIT	54486.5	5.044	93781.5	6.257	81449.5	9.549
EPA SAMPLE NO.						
WB-301-SW	80908	5.55	148484	6.76	126066	10.06
TB-10102024	87697	5.55	161174	6.76	134772	10.06
VX1015WBL01	94368	5.55	170879	6.76	144567	10.06
VX1015WBS01	98951	5.55	177022	6.76	148487	10.06
VX1015WBSD01	94177	5.54	163248	6.76	138792	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.