

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4397 SAS No.: P4397 SDG NO.: P4397
 Lab File ID: VX043389.D Date Analyzed: 10/15/2024
 Instrument ID: MSVOA_X Time Analyzed: 08:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	75623	12.018				
UPPER LIMIT	151246	12.518				
LOWER LIMIT	37811.5	11.518				
EPA SAMPLE NO.						
WB-301-SW	50389	12.02				
TB-10102024	57811	12.02				
VX1015WBL01	64756	12.02				
VX1015WBS01	66651	12.02				
VX1015WBSD01	64769	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.