

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P4423

Project: Monthly 2024

Run No.: LB132966

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB132966BS							
BOD5	mg/L	198	195		98	1	84.6-115.4	10/16/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P4423

Project: Monthly 2024

Run No.: LB132972

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB132972BS							
TSS	mg/L	550	532		97	1	90-110	10/17/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P4423

Project: Monthly 2024

Run No.: LB132973

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB132973BS							
TPH	mg/L	20.0	16.7		84	1	78-114	10/17/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P4423

Project: Monthly 2024

Run No.: LB132973

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB132973BSD							
TPH	mg/L	20.0	16.9		84	1	78-114	10/17/2024