



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms **SDG No.:** P4423
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4423 **SAS No.:** P4423
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	93	40 - 160	CV	10/18/2024	12:15	LB132999
CRI01	Cadmium	5.58	6.0	93	40 - 160	P	10/24/2024	13:38	LB133110
	Copper	22.1	20.0	110	40 - 160	P	10/24/2024	13:38	LB133110
	Lead	11.7	12.0	97	40 - 160	P	10/24/2024	13:38	LB133110
	Nickel	38.8	40.0	97	40 - 160	P	10/24/2024	13:38	LB133110
	Zinc	44.4	40.0	111	40 - 160	P	10/24/2024	13:38	LB133110