

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Aramark Uniforms **SDG No.:** P4423
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4423 **SAS No.:** P4423
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	3.37	1.0	337	-5	7	10/24/2024	13:43	LB133110
	Copper	8.59	2.0	430	-18	22	10/24/2024	13:43	LB133110
	Lead	9.27			-12	12	10/24/2024	13:43	LB133110
	Nickel	2.59	2.0	130	-38	42	10/24/2024	13:43	LB133110
	Zinc	5.89			-40	40	10/24/2024	13:43	LB133110
ICSAB01	Cadmium	981	972	101	826	1120	10/24/2024	13:54	LB133110
	Copper	518	511	101	434	588	10/24/2024	13:54	LB133110
	Lead	57.7	49.0	118	37	61	10/24/2024	13:54	LB133110
	Nickel	1040	954	109	810	1100	10/24/2024	13:54	LB133110
	Zinc	918	952	96	809	1095	10/24/2024	13:54	LB133110