

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Aramark Uniforms **SDG No.:** P4423
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** P4423 **SAS No.:** P4423
Matrix: Water **Level:** LOW **Client ID:** WB-303-SWA
Sample ID: P4460-06 **Spiked ID:** P4460-06A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Zinc	ug/L	75 - 125	161		19.8	J	100	141		P