



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: XRDS Recycling **SDG No.:** P4424
Contract: XRDS01 **Lab Code:** CHEM **Case No.:** P4424 **SAS No.:** P4424
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Silver	10.1	10.0	101	40 - 160	P	10/23/2024	15:22	LB133086
CRI01	Silver	10.6	10.0	106	40 - 160	P	10/24/2024	13:38	LB133110