

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: XRDS Recycling **SDG No.:** P4424
Contract: XRDS01 **Lab Code:** CHEM **Case No.:** P4424 **SAS No.:** P4424
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Silver	0.40			-10	10	10/23/2024	15:26	LB133086
ICSAB01	Silver	206	201	102	170	232	10/23/2024	15:32	LB133086
ICSA01	Silver	0.37			-10	10	10/24/2024	13:43	LB133110
ICSAB01	Silver	230	201	114	170	232	10/24/2024	13:54	LB133110