

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: P4428

SDG NO.: P4428

Lab File ID: BN034683.D

DFTPP Injection Date: 10/24/2024

Instrument ID: BNA_N

DFTPP Injection Time: 07:59

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	66.1
68	Less than 2.0% of mass 69	0.9 (1.7) 1
69	Mass 69 relative abundance	54.4
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	60.3
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	20.6
365	Greater than 1% of mass 198	2.5
441	Present, but less than mass 443	7.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	9.1 (18.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN034684.D	10/24/2024	09:11
SSTDICC0.2	SSTDICC0.2	BN034685.D	10/24/2024	09:47
SSTDICCC0.4	SSTDICCC0.4	BN034686.D	10/24/2024	10:23
SSTDICC0.8	SSTDICC0.8	BN034687.D	10/24/2024	10:59
SSTDICC1.6	SSTDICC1.6	BN034688.D	10/24/2024	11:35
SSTDICC3.2	SSTDICC3.2	BN034689.D	10/24/2024	12:11
SSTDICC5.0	SSTDICC5.0	BN034690.D	10/24/2024	12:48
PB164229BL	PB164229BL	BN034692.D	10/24/2024	14:50
RW7-SP100-20241016	P4428-01	BN034693.D	10/24/2024	15:28
PB164229BS	PB164229BS	BN034694.D	10/24/2024	16:05
RW7-SP302-20241016	P4428-03	BN034695.D	10/24/2024	16:41
RW7-SP303-20241016	P4428-04	BN034696.D	10/24/2024	17:18
SSTDCCC0.4EC	SSTDCCC0.4	BN034699.D	10/24/2024	19:06