


LD133003

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.074	0.074	360	0.000	360					7
0.2	0.200	0.210	0.010	1037	0.0 %	1037					5
2.5	2.500	2.261	-0.239	11242	0.0 %	11242					10
5.0	5.000	5.138	0.138	25550	0.0 %	25550					3
7.5	7.500	7.657	0.157	38076	0.0 %	38076					2
10.0	10.000	9.860	-0.140	49038	0.0 %	49038					1

LB133003 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Oct 2024 13:50:41

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	360	-	0.0000	7471B	PPB	18 Oct 2024 13:59:54	S	Std
0.2 - 1	50.2	1037	-	0.2000	7471B	PPB	18 Oct 2024 14:09:28	S	Std
2.5 - 1	52.5	11242	-	2.5000	7471B	PPB	18 Oct 2024 14:11:45	S	Std
5.0 - 1	55	25550	-	5.0000	7471B	PPB	18 Oct 2024 14:16:40	S	Std
7.5 - 1	57.5	38076	-	7.5000	7471B	PPB	18 Oct 2024 14:21:58	S	Std
10.0 - 1	510	49038	-	10.0000	7471B	PPB	18 Oct 2024 14:26:58	S	Std
ICV44 - 1	ICV44	19607	3.9433	-	7471B	PPB	18 Oct 2024 14:30:03	U	SMPL
ICB44 - 1	ICB44	-338	-0.0667	-	7471B	PPB	18 Oct 2024 14:32:18	U	SMPL
CCV42 - 1	CCV42	25592	5.1466	-	7471B	PPB	18 Oct 2024 14:34:36	U	SMPL
CCB42 - 1	CCB42	-270	-0.0530	-	7471B	PPB	18 Oct 2024 14:36:50	U	SMPL
CRA - 1	CRA	1072	0.2168	-	7471B	PPB	18 Oct 2024 14:39:08	U	SMPL
HighStd - 1	HighStd	48238	9.6996	-	7471B	PPB	18 Oct 2024 14:41:23	U	SMPL
ChkStd - 1	ChkStd	35411	7.1207	-	7471B	PPB	18 Oct 2024 14:43:38	U	SMPL
PB164258BL - 1	PBS	-381	-0.0753	-	7471B	PPB	18 Oct 2024 14:45:59	U	SMPL
PB164258BS - 1	LCSS	19815	3.9851	-	7471B	PPB	18 Oct 2024 14:50:36	U	SMPL
P4425-01 - 1	TP-1	2421	0.4880	-	7471B	PPB	18 Oct 2024 14:57:51	U	SMPL
P4425-01DUP - 1	TP-1DUP	2106	0.4247	-	7471B	PPB	18 Oct 2024 15:00:07	U	SMPL
P4425-01MS - 1	TP-1MS	23761	4.7785	-	7471B	PPB	18 Oct 2024 15:02:25	U	SMPL
P4425-01MSD - 1	TP-1MSD	25552	5.1385	-	7471B	PPB	18 Oct 2024 15:04:42	U	SMPL
P4425-03 - 1	TP-2	4154	0.8364	-	7471B	PPB	18 Oct 2024 15:06:59	U	SMPL
CCV43 - 1	CCV43	25324	5.0927	-	7471B	PPB	18 Oct 2024 15:09:18	U	SMPL
CCB43 - 1	CCB43	-296	-0.0582	-	7471B	PPB	18 Oct 2024 15:11:35	U	SMPL
P4425-05 - 1	TP-3	3740	0.7532	-	7471B	PPB	18 Oct 2024 15:13:54	U	SMPL
P4425-07 - 1	TP-4	562	0.1143	-	7471B	PPB	18 Oct 2024 15:16:09	U	SMPL
P4425-09 - 1	TP-5	1074	0.2172	-	7471B	PPB	18 Oct 2024 15:18:25	U	SMPL
P4431-01 - 1	72-11934	1589	0.3207	-	7471B	PPB	18 Oct 2024 15:20:41	U	SMPL
P4432-01 - 1	LOT-4078A	-18	-0.0023	-	7471B	PPB	18 Oct 2024 15:22:56	U	SMPL
P4425-01LX5 - 1		346	0.0708	-	7471B	PPB	18 Oct 2024 15:25:13	U	SMPL
P4425-01A - 1		25442	5.1164	-	7471B	PPB	18 Oct 2024 15:27:29	U	SMPL
CCV44 - 1	CCV44	26409	5.3108	-	7471B	PPB	18 Oct 2024 15:29:45	U	SMPL
CCB44 - 1	CCB44	-256	-0.0502	-	7471B	PPB	18 Oct 2024 15:32:05	U	SMPL