

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133003

Review By	jaswal	Review On	10/18/2024 8:40:37 PM
Supervise By	mohan	Supervise On	10/18/2024 8:42:40 PM

STD. NAME	STD REF.#
ICAL Standard	MP82803,MP82805,MP82806,MP82807,MP82808,MP82809
ICV Standard	MP82810
CCV Standard	MP82812
ICSA Standard	
CRI Standard	MP82814
LCS Standard	
Chk Standard	MP82811,MP82813,MP82815,MP820

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/18/24 13:59		Mohan	OK
2	S0.2	S0.2	CAL2	10/18/24 14:09		Mohan	OK
3	S2.5	S2.5	CAL3	10/18/24 14:11		Mohan	OK
4	S5	S5	CAL4	10/18/24 14:16		Mohan	OK
5	S7.5	S7.5	CAL5	10/18/24 14:21		Mohan	OK
6	S10	S10	CAL6	10/18/24 14:26		Mohan	OK
7	ICV44	ICV44	ICV	10/18/24 14:30		Mohan	OK
8	ICB44	ICB44	ICB	10/18/24 14:32		Mohan	OK
9	CCV42	CCV42	CCV	10/18/24 14:34		Mohan	OK
10	CCB42	CCB42	CCB	10/18/24 14:36		Mohan	OK
11	CRA	CRA	CRDL	10/18/24 14:39		Mohan	OK
12	HighStd	HighStd	HIGH STD	10/18/24 14:41		Mohan	OK
13	ChkStd	ChkStd	SAM	10/18/24 14:43		Mohan	OK
14	PB164258BL	PB164258BL	MB	10/18/24 14:45		Mohan	OK
15	PB164258BS	PB164258BS	LCS	10/18/24 14:50		Mohan	OK
16	P4425-01	TP-1	SAM	10/18/24 14:57		Mohan	OK
17	P4425-01DUP	TP-1DUP	DUP	10/18/24 15:00		Mohan	OK
18	P4425-01MS	TP-1MS	MS	10/18/24 15:02		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB133003

Review By	jaswal	Review On	10/18/2024 8:40:37 PM
Supervise By	mohan	Supervise On	10/18/2024 8:42:40 PM

STD. NAME	STD REF.#
ICAL Standard	MP82803,MP82805,MP82806,MP82807,MP82808,MP82809
ICV Standard	MP82810
CCV Standard	MP82812
ICSA Standard	
CRI Standard	MP82814
LCS Standard	
Chk Standard	MP82811,MP82813,MP82815,MP820

19	P4425-01MSD	TP-1MSD	MSD	10/18/24 15:04		Mohan	OK
20	P4425-03	TP-2	SAM	10/18/24 15:06		Mohan	OK
21	CCV43	CCV43	CCV	10/18/24 15:09		Mohan	OK
22	CCB43	CCB43	CCB	10/18/24 15:11		Mohan	OK
23	P4425-05	TP-3	SAM	10/18/24 15:13		Mohan	OK
24	P4425-07	TP-4	SAM	10/18/24 15:16		Mohan	OK
25	P4425-09	TP-5	SAM	10/18/24 15:18		Mohan	OK
26	P4431-01	72-11934	SAM	10/18/24 15:20		Mohan	OK
27	P4432-01	LOT-4078A	SAM	10/18/24 15:22		Mohan	OK
28	P4425-01L	TP-1L	SD	10/18/24 15:25		Mohan	OK
29	P4425-01A	TP-1A	PS	10/18/24 15:27		Mohan	OK
30	CCV44	CCV44	CCV	10/18/24 15:29		Mohan	OK
31	CCB44	CCB44	CCB	10/18/24 15:32		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133039

Review By	kareem	Review On	10/25/2024 2:57:15 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:06:00 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/21/24 18:11		Kareem	OK
2	S1	S1	CAL2	10/21/24 18:16		Kareem	OK
3	S2	S2	CAL3	10/21/24 18:20		Kareem	OK
4	S3	S3	CAL4	10/21/24 18:24		Kareem	OK
5	S4	S4	CAL5	10/21/24 18:28		Kareem	OK
6	S5	S5	CAL6	10/21/24 18:32		Kareem	OK
7	ICV01	ICV01	ICV	10/21/24 18:38		Kareem	OK
8	LLICV01	LLICV01	LLICV	10/21/24 18:42		Kareem	OK
9	ICB01	ICB01	ICB	10/21/24 18:46		Kareem	OK
10	CRI01	CRI01	CRDL	10/21/24 18:51		Kareem	OK
11	ICSA01	ICSA01	ICSA	10/21/24 18:55		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	10/21/24 19:00		Kareem	OK
13	CCV01	CCV01	CCV	10/21/24 19:04		Kareem	OK
14	CCB01	CCB01	CCB	10/21/24 19:14		Kareem	OK
15	P4417-10	HOPPER-A-WATER	SAM	10/21/24 19:18		Kareem	OK
16	P4417-12	HOPPER-B-WATER	SAM	10/21/24 19:22		Kareem	OK
17	P4425-01	TP-1	SAM	10/21/24 19:27		Kareem	OK
18	P4425-03	TP-2	SAM	10/21/24 19:31		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133039

Review By	kareem	Review On	10/25/2024 2:57:15 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:06:00 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712		
ICV Standard	mp82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82712		
LCS Standard			
Chk Standard	MP82491 MP82492		

19	P4425-05	TP-3	SAM	10/21/24 19:35		Kareem	OK
20	P4425-07	TP-4	SAM	10/21/24 19:39		Kareem	OK
21	P4425-09	TP-5	SAM	10/21/24 19:43		Kareem	OK
22	P4417-02	HOPPER-A-SOLID	SAM	10/21/24 19:47	Cr high	Kareem	Dilution
23	P4417-05	HOPPER-B-SOLID	SAM	10/21/24 19:51	Cr high	Kareem	Dilution
24	P4417-08	HOPPER-C-SOLID	SAM	10/21/24 19:55	Cr high	Kareem	Dilution
25	CCV02	CCV02	CCV	10/21/24 19:59		Kareem	OK
26	CCB02	CCB02	CCB	10/21/24 20:04		Kareem	OK
27	P4417-03	HOPPER-A-SOLID	SAM	10/21/24 20:08		Kareem	OK
28	P4417-06	HOPPER-B-SOLID	SAM	10/21/24 20:12		Kareem	OK
29	P4417-09	HOPPER-C-SOLID	SAM	10/21/24 20:17		Kareem	OK
30	P4419-04	IB-1	SAM	10/21/24 20:21		Kareem	OK
31	P4425-02	TP-1	SAM	10/21/24 20:25		Kareem	OK
32	P4425-04	TP-2	SAM	10/21/24 20:30		Kareem	OK
33	P4425-06	TP-3	SAM	10/21/24 20:34		Kareem	OK
34	P4425-08	TP-4	SAM	10/21/24 20:39		Kareem	OK
35	P4425-10	TP-5	SAM	10/21/24 20:43		Kareem	OK
36	P4417-11	HOPPER-A-WATER	SAM	10/21/24 20:48		Kareem	OK
37	CCV03	CCV03	CCV	10/21/24 20:52		Kareem	OK
38	CCB03	CCB03	CCB	10/21/24 20:56		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133039

Review By	kareem	Review On	10/25/2024 2:57:15 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:06:00 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

39	P4417-02DL	HOPPER-A-SOLIDDL	SAM	10/21/24 21:01	5x for Cr	Kareem	Confirms
40	P4417-05DL	HOPPER-B-SOLIDDL	SAM	10/21/24 21:05	5x for Cr	Kareem	Confirms
41	P4417-08DL	HOPPER-C-SOLIDDL	SAM	10/21/24 21:09	5x for Cr	Kareem	Confirms
42	P4417-13	HOPPER-B-WATER	SAM	10/21/24 21:13		Kareem	OK
43	CCV04	CCV04	CCV	10/21/24 21:17		Kareem	OK
44	CCB04	CCB04	CCB	10/21/24 21:22		Kareem	OK
45	LR1	LR1	HIGH STD	10/21/24 21:27		Kareem	OK
46	LR2	LR2	HIGH STD	10/21/24 21:32		Kareem	OK
47	P4425-09DUP	TP-5DUP	DUP	10/21/24 21:36		Kareem	OK
48	P4425-09L	TP-5L	SD	10/21/24 21:40		Kareem	OK
49	P4425-09MS	TP-5MS	MS	10/21/24 21:45	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
50	P4425-09MSD	TP-5MSD	MSD	10/21/24 21:49	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
51	P4425-09A	TP-5A	PS	10/21/24 21:53	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
52	P4417-05DUP	HOPPER-B-SOLIDDU	DUP	10/21/24 21:56	Cr high	Kareem	Dilution
53	P4417-05L	HOPPER-B-SOLIDL	SD	10/21/24 22:01		Kareem	OK
54	P4417-05MS	HOPPER-B-SOLIDMS	MS	10/21/24 22:05	5x for Cr	Kareem	Dilution
55	CCV05	CCV05	CCV	10/21/24 22:09		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133039

Review By	kareem	Review On	10/25/2024 2:57:15 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:06:00 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

56	CCB05	CCB05	CCB	10/21/24 22:13		Kareem	OK
57	P4417-05MSD	HOPPER-B-SOLIDMS	MSD	10/21/24 22:17	5x for Cr	Kareem	Dilution
58	P4417-05A	HOPPER-B-SOLIDA	PS	10/21/24 22:21	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
59	P4417-13DUP	HOPPER-B-WATERD	DUP	10/21/24 22:25		Kareem	OK
60	P4417-13L	HOPPER-B-WATERL	SD	10/21/24 22:29		Kareem	OK
61	P4417-13MS	HOPPER-B-WATERM	MS	10/21/24 22:34	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
62	P4417-13MSD	HOPPER-B-WATERM	MSD	10/21/24 22:38	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
63	P4417-13A	HOPPER-B-WATERA	PS	10/21/24 22:42	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
64	P4417-05DUPDL	HOPPER-B-SOLIDDU	DUP	10/21/24 22:46	5x for Cr	Kareem	Confirms
65	P4417-05LDL	HOPPER-B-SOLIDLD	SD	10/21/24 22:50	Not Required	Kareem	Not Ok
66	P4417-05MSDL	HOPPER-B-SOLIDMS	MS	10/21/24 22:54	5x for Cr	Kareem	Confirms
67	CCV06	CCV06	CCV	10/21/24 22:58		Kareem	OK
68	CCB06	CCB06	CCB	10/21/24 23:02		Kareem	OK
69	P4417-05MSDRE	HOPPER-B-SOLIDMS	MSD	10/21/24 23:07	NOT USE	Kareem	Not Ok
70	P4417-05MSDDL	HOPPER-B-SOLIDMS	MSD	10/21/24 23:11	5x for Cr	Kareem	Confirms
71	P4417-05ADL	HOPPER-B-SOLIDA	PS	10/21/24 23:16	Not Required	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133039

Review By	kareem	Review On	10/25/2024 2:57:15 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:06:00 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712		
ICV Standard	mp82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82712		
LCS Standard			
Chk Standard	MP82491 MP82492		

72	P4417-12DUP	HOPPER-B-WATERD	DUP	10/21/24 23:21		Kareem	OK
73	P4417-12L	HOPPER-B-WATERL	SD	10/21/24 23:25		Kareem	OK
74	P4417-12MS	HOPPER-B-WATERM	MS	10/21/24 23:29	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
75	P4417-12MSD	HOPPER-B-WATERM	MSD	10/21/24 23:33	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
76	P4417-12A	HOPPER-B-WATERA	PS	10/21/24 23:37	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
77	CCV07	CCV07	CCV	10/21/24 23:42		Kareem	OK
78	CCB07	CCB07	CCB	10/21/24 23:46		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133061

Review By	kareem	Review On	10/25/2024 3:01:56 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712		
ICV Standard	mp82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82712		
LCS Standard			
Chk Standard	MP82491 MP82492		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/22/24 14:36		Kareem	OK
2	S1	S1	CAL2	10/22/24 14:40		Kareem	OK
3	S2	S2	CAL3	10/22/24 14:45		Kareem	OK
4	S3	S3	CAL4	10/22/24 14:49		Kareem	OK
5	S4	S4	CAL5	10/22/24 14:53		Kareem	OK
6	S5	S5	CAL6	10/22/24 14:57		Kareem	OK
7	ICV01	ICV01	ICV	10/22/24 15:01	ICV fail for Mg,V (200.7)	Kareem	OK
8	LLICV01	LLICV01	LLICV	10/22/24 15:16		Kareem	OK
9	ICB01	ICB01	ICB	10/22/24 15:20		Kareem	OK
10	CRI01	CRI01	CRDL	10/22/24 15:24		Kareem	OK
11	ICSA01	ICSA01	ICSA	10/22/24 15:29		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	10/22/24 15:33		Kareem	OK
13	ICSADL	ICSADL	ICSA	10/22/24 15:38		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	10/22/24 15:42		Kareem	OK
15	CCV01	CCV01	CCV	10/22/24 15:52		Kareem	OK
16	CCB01	CCB01	CCB	10/22/24 15:56		Kareem	OK
17	P4347-01	EFFLUENT-DAY-1-MI	SAM	10/22/24 16:01	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
18	P4426-02	PAD-1	SAM	10/22/24 16:05		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133061

Review By	kareem	Review On	10/25/2024 3:01:56 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712		
ICV Standard	mp82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82712		
LCS Standard			
Chk Standard	MP82491 MP82492		

19	P4430-01	VNJ-209	SAM	10/22/24 16:10		Kareem	OK
20	P4455-01	SU-4-101824	SAM	10/22/24 16:26		Kareem	OK
21	CCV02	CCV02	CCV	10/22/24 16:30		Kareem	OK
22	CCB02	CCB02	CCB	10/22/24 16:34		Kareem	OK
23	P4426-04	PAD-2	SAM	10/22/24 16:39		Kareem	OK
24	P4426-06	PAD-3	SAM	10/22/24 16:44		Kareem	OK
25	P4426-08	PAD-4	SAM	10/22/24 16:48		Kareem	OK
26	P4426-10	PAD-5	SAM	10/22/24 16:53		Kareem	OK
27	P4426-12	PAD-6	SAM	10/22/24 16:57		Kareem	OK
28	P4426-14	PAD-7	SAM	10/22/24 17:02		Kareem	OK
29	P4426-16	PAD-8	SAM	10/22/24 17:06		Kareem	OK
30	P4426-18	PAD-9	SAM	10/22/24 17:11		Kareem	OK
31	P4426-20	PAD-10	SAM	10/22/24 17:16		Kareem	OK
32	P4426-22	PAD-11	SAM	10/22/24 17:20		Kareem	OK
33	CCV03	CCV03	CCV	10/22/24 17:25		Kareem	OK
34	CCB03	CCB03	CCB	10/22/24 17:29		Kareem	OK
35	P4426-24	PAD-12	SAM	10/22/24 17:37		Kareem	OK
36	P4426-26	PAD-13	SAM	10/22/24 17:42		Kareem	OK
37	P4426-28	PAD-14	SAM	10/22/24 17:47		Kareem	OK
38	P4426-30	PAD-15	SAM	10/22/24 17:51		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133061

Review By	kareem	Review On	10/25/2024 3:01:56 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712		
ICV Standard	mp82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82712		
LCS Standard			
Chk Standard	MP82491 MP82492		

39	P4426-32	PAD-16	SAM	10/22/24 17:56		Kareem	OK
40	P4426-34	PAD-17	SAM	10/22/24 18:00		Kareem	OK
41	P4426-36	TRANSFORMER-PAD	SAM	10/22/24 18:05		Kareem	OK
42	P4426-38	WALL	SAM	10/22/24 18:10		Kareem	OK
43	P4426-40	CURBS-AND-OTHER	SAM	10/22/24 18:14		Kareem	OK
44	CCV04	CCV04	CCV	10/22/24 18:19		Kareem	OK
45	CCB04	CCB04	CCB	10/22/24 18:23		Kareem	OK
46	P4426-01	PAD-1	SAM	10/22/24 18:28		Kareem	OK
47	P4426-03	PAD-2	SAM	10/22/24 18:33		Kareem	OK
48	P4426-05	PAD-3	SAM	10/22/24 18:37		Kareem	OK
49	P4426-07	PAD-4	SAM	10/22/24 18:42		Kareem	OK
50	P4426-09	PAD-5	SAM	10/22/24 18:46		Kareem	OK
51	P4426-11	PAD-6	SAM	10/22/24 18:51		Kareem	OK
52	P4426-13	PAD-7	SAM	10/22/24 18:55		Kareem	OK
53	P4426-15	PAD-8	SAM	10/22/24 19:00		Kareem	OK
54	P4426-17	PAD-9	SAM	10/22/24 19:04		Kareem	OK
55	P4426-21	PAD-11	SAM	10/22/24 19:09		Kareem	OK
56	CCV05	CCV05	CCV	10/22/24 19:13		Kareem	OK
57	CCB05	CCB05	CCB	10/22/24 19:17		Kareem	OK
58	P4426-23	PAD-12	SAM	10/22/24 19:23		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133061

Review By	kareem	Review On	10/25/2024 3:01:56 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:11 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

59	P4426-25	PAD-13	SAM	10/22/24 19:27		Kareem	OK
60	P4426-27	PAD-14	SAM	10/22/24 19:32		Kareem	OK
61	P4426-29	PAD-15	SAM	10/22/24 19:36		Kareem	OK
62	P4426-31	PAD-16	SAM	10/22/24 19:40		Kareem	OK
63	P4426-33	PAD-17	SAM	10/22/24 19:45		Kareem	OK
64	P4426-35	TRANSFORMER-PAD	SAM	10/22/24 19:49		Kareem	OK
65	P4426-37	WALL	SAM	10/22/24 19:54		Kareem	OK
66	P4426-39	CURBS-AND-OTHER	SAM	10/22/24 19:58		Kareem	OK
67	P4347-01DUP	EFFLUENT-DAY-1-MI	DUP	10/22/24 20:02	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
68	CCV06	CCV06	CCV	10/22/24 20:24		Kareem	OK
69	CCB06	CCB06	CCB	10/22/24 20:29		Kareem	OK
70	P4347-01L	EFFLUENT-DAY-1-MI	SD	10/22/24 20:33	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
71	P4347-01MS	EFFLUENT-DAY-1-MI	MS	10/22/24 20:38	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
72	P4347-01MSD	EFFLUENT-DAY-1-MI	MSD	10/22/24 20:42	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
73	P4347-01A	EFFLUENT-DAY-1-MI	PS	10/22/24 20:46	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
74	P4426-19	PAD-10	SAM	10/22/24 20:50		Kareem	OK
75	P4426-39DUP	CURBS-AND-OTHER	DUP	10/22/24 20:54		Kareem	OK
76	P4426-39L	CURBS-AND-OTHER	SD	10/22/24 20:58		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133061

Review By	kareem	Review On	10/25/2024 3:01:56 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:11 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

77	P4426-39MS	CURBS-AND-OTHER	MS	10/22/24 21:03	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
78	P4426-39MSD	CURBS-AND-OTHER	MSD	10/22/24 21:07	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
79	P4426-39A	CURBS-AND-OTHER	PS	10/22/24 21:11	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
80	CCV07	CCV07	CCV	10/22/24 21:16		Kareem	OK
81	CCB07	CCB07	CCB	10/22/24 21:20		Kareem	OK
82	P4463-01	001-WILLETS-PT-BL	SAM	10/22/24 21:32		Kareem	OK
83	P4463-02	002-35TH-AVE(OCT)	SAM	10/22/24 21:40		Kareem	OK
84	P4464-01	001-WILLETS-PT-BL	SAM	10/22/24 21:44		Kareem	OK
85	P4464-02	002-35TH-AVE(SEP)	SAM	10/22/24 21:48		Kareem	OK
86	PB164302BL	PB164302BL	MB	10/22/24 21:53	LCS fail for Zn	Kareem	Not Ok
87	PB164302BS	PB164302BS	LCS	10/22/24 21:57	LCS fail for Zn	Kareem	Not Ok
88	P4417-10	HOPPER-A-WATER	SAM	10/22/24 22:11		Kareem	OK
89	P4419-01	IB-1	SAM	10/22/24 22:15	Cr high	Kareem	Dilution
90	LR1	LR1	HIGH STD	10/22/24 22:20		Kareem	OK
91	LR2	LR2	HIGH STD	10/22/24 22:25		Kareem	OK
92	CCV08	CCV08	CCV	10/22/24 22:29		Kareem	OK
93	CCB08	CCB08	CCB	10/22/24 22:33		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133061

Review By	kareem	Review On	10/25/2024 3:01:56 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:11 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

94	P4419-01DL	IB-1DL	SAM	10/22/24 22:37	5x for Cr	Kareem	Confirms
95	P4426-40DUP	CURBS-AND-OTHER	DUP	10/22/24 22:41		Kareem	OK
96	P4426-40L	CURBS-AND-OTHER	SD	10/22/24 22:46		Kareem	OK
97	P4426-40MS	CURBS-AND-OTHER	MS	10/22/24 22:50	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
98	P4426-40MSD	CURBS-AND-OTHER	MSD	10/22/24 22:54	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
99	P4426-40A	CURBS-AND-OTHER	PS	10/22/24 22:59	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
100	PB164218TB	PB164218TB	MB	10/22/24 23:03		Kareem	OK
101	PB164233BL	PB164233BL	MB	10/22/24 23:07		Kareem	OK
102	PB164233BS	PB164233BS	LCS	10/22/24 23:12	0.1ML OF M6010 AND M6001 WERE ADDED TO 10ML OF SAMPLE	Kareem	OK
103	P4432-01	LOT-4078A	SAM	10/22/24 23:16		Kareem	OK
104	CCV09	CCV09	CCV	10/22/24 23:20		Kareem	OK
105	CCB09	CCB09	CCB	10/22/24 23:24		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712		
ICV Standard	mp82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82712		
LCS Standard			
Chk Standard	MP82491 MP82492		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/23/24 14:24		Kareem	OK
2	S1	S1	CAL2	10/23/24 14:28		Kareem	OK
3	S2	S2	CAL3	10/23/24 14:32		Kareem	OK
4	S3	S3	CAL4	10/23/24 14:37		Kareem	OK
5	S4	S4	CAL5	10/23/24 14:41		Kareem	OK
6	S5	S5	CAL6	10/23/24 14:45		Kareem	OK
7	ICV01	ICV01	ICV	10/23/24 14:59		Kareem	OK
8	LLICV01	LLICV01	LLICV	10/23/24 15:11		Kareem	OK
9	ICB01	ICB01	ICB	10/23/24 15:16		Kareem	OK
10	CRI01	CRI01	CRDL	10/23/24 15:22		Kareem	OK
11	ICSA01	ICSA01	ICSA	10/23/24 15:26		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	10/23/24 15:32		Kareem	OK
13	ICSADL	ICSADL	ICSA	10/23/24 15:36		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	10/23/24 15:40		Kareem	OK
15	CCV01	CCV01	CCV	10/23/24 15:57		Kareem	OK
16	CCB01	CCB01	CCB	10/23/24 16:02		Kareem	OK
17	P4458-02	280517	SAM	10/23/24 16:06		Kareem	OK
18	P4458-01	280517	SAM	10/23/24 16:10		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

19	P4443-05	OG-315-HR-502-COM	SAM	10/23/24 16:15		Kareem	OK
20	P4443-10	OG-315-HR-502-COM	SAM	10/23/24 16:19		Kareem	OK
21	P4443-01	OG-315-HR-502-COM	SAM	10/23/24 16:23		Kareem	OK
22	P4443-06	OG-315-HR-502-COM	SAM	10/23/24 16:28		Kareem	OK
23	P4452-01	ETGI-285	SAM	10/23/24 16:32	Ca,Cr high	Kareem	Dilution
24	P4452-01DL	ETGI-285DL	SAM	10/23/24 16:44	5x for Ca,Cr	Kareem	Confirms
25	CCV02	CCV02	CCV	10/23/24 16:48		Kareem	OK
26	CCB02	CCB02	CCB	10/23/24 16:52		Kareem	OK
27	P4397-06	WB-301-BOT	SAM	10/23/24 16:59		Kareem	OK
28	P4460-04	WB-303-BOT	SAM	10/23/24 17:03		Kareem	OK
29	P4460-04DUP	WB-303-BOTDUP	DUP	10/23/24 17:08		Kareem	OK
30	P4460-04L	WB-303-BOTL	SD	10/23/24 17:12		Kareem	OK
31	P4460-04MS	WB-303-BOTMS	MS	10/23/24 17:17	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
32	P4460-04MSD	WB-303-BOTMSD	MSD	10/23/24 17:21	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
33	P4460-04A	WB-303-BOTA	PS	10/23/24 17:25	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
34	PB164261TB	PB164261TB	MB	10/23/24 17:29		Kareem	OK
35	PB164298BL	PB164298BL	MB	10/23/24 17:34		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

36	PB164298BS	PB164298BS	LCS	10/23/24 17:38	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
37	CCV03	CCV03	CCV	10/23/24 17:42		Kareem	OK
38	CCB03	CCB03	CCB	10/23/24 17:46		Kareem	OK
39	P4456-01	PAD-10182024	SAM	10/23/24 17:51		Kareem	OK
40	P4460-02	WB-303-TOP	SAM	10/23/24 17:55		Kareem	OK
41	P4460-03	WB-303-BOT	SAM	10/23/24 17:59		Kareem	OK
42	P4460-03DUP	WB-303-BOTDUP	DUP	10/23/24 18:04		Kareem	OK
43	P4460-03L	WB-303-BOTL	SD	10/23/24 18:08		Kareem	OK
44	P4460-03MS	WB-303-BOTMS	MS	10/23/24 18:12		Kareem	OK
45	P4460-03MSD	WB-303-BOTMSD	MSD	10/23/24 18:16		Kareem	OK
46	P4460-03A	WB-303-BOTA	PS	10/23/24 18:20		Kareem	OK
47	PB164289BL	PB164289BL	MB	10/23/24 18:24	Fail for Al	Kareem	Not Ok
48	PB164289BS	PB164289BS	LCS	10/23/24 18:28	Fail for Al,Sb,As,Ba,Be,Cd,Pb,Mg, Mn,Se,Tl,V	Kareem	Not Ok
49	CCV04	CCV04	CCV	10/23/24 18:32		Kareem	OK
50	CCB04	CCB04	CCB	10/23/24 18:36		Kareem	OK
51	PB164244BL	PB164244BL	MB	10/23/24 18:41		Kareem	OK
52	PB164244BS	PB164244BS	LCS	10/23/24 18:45	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712		
ICV Standard	mp82485		
CCV Standard	MP82488		
ICSA Standard	MP82486 MP82487		
CRI Standard	MP82712		
LCS Standard			
Chk Standard	MP82491 MP82492		

53	PB164195TB	PB164195TB	MB	10/23/24 18:49		Kareem	OK
54	PB164197TB	PB164197TB	MB	10/23/24 18:54		Kareem	OK
55	PB164248BL	PB164248BL	MB	10/23/24 18:58		Kareem	OK
56	PB164248BS	PB164248BS	LCS	10/23/24 19:03	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
57	PB164174BL	PB164174BL	MB	10/23/24 19:07		Kareem	OK
58	PB164174BS	PB164174BS	LCS	10/23/24 19:11	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
59	PB164221BL	PB164221BL	MB	10/23/24 19:15		Kareem	OK
60	PB164221BS	PB164221BS	LCS	10/23/24 19:19	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
61	CCV05	CCV05	CCV	10/23/24 19:23		Kareem	OK
62	CCB05	CCB05	CCB	10/23/24 19:28		Kareem	OK
63	PB164222BL	PB164222BL	MB	10/23/24 19:32		Kareem	OK
64	PB164222BS	PB164222BS	LCS	10/23/24 19:36	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
65	PB164220BL	PB164220BL	MB	10/23/24 19:40		Kareem	OK
66	PB164220BS	PB164220BS	LCS	10/23/24 19:45	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
67	P4424-01	1-N-1	SAM	10/23/24 19:49		Kareem	OK
68	P4424-01DUP	1-N-1DUP	DUP	10/23/24 19:53		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

69	P4424-01L	1-N-1L	SD	10/23/24 19:58		Kareem	OK
70	P4424-01MS	1-N-1MS	MS	10/23/24 20:02	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
71	P4424-01MSD	1-N-1MSD	MSD	10/23/24 20:06	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
72	P4424-01A	1-N-1A	PS	10/23/24 20:11	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
73	CCV06	CCV06	CCV	10/23/24 20:15		Kareem	OK
74	CCB06	CCB06	CCB	10/23/24 20:21		Kareem	OK
75	P4424-02	1-S-1	SAM	10/23/24 20:25		Kareem	OK
76	P4424-03	1-B-1	SAM	10/23/24 20:30		Kareem	OK
77	P4397-04DL	WB-301-SWDL	SAM	10/23/24 20:47	5x for Na	Kareem	Confirms
78	P4397-04DUPDL	WB-301-SWDUPDL	DUP	10/23/24 20:51	5x for Na	Kareem	Confirms
79	P4397-04LDL	WB-301-SWLDL	SD	10/23/24 20:56	Not Required	Kareem	Not Ok
80	P4397-04MSDL	WB-301-SWMSDL	MS	10/23/24 21:00	5x for Na	Kareem	Confirms
81	P4397-04MSDDL	WB-301-SWMSDDL	MSD	10/23/24 21:05	5x for Na	Kareem	Confirms
82	LR1	LR1	HIGH STD	10/23/24 21:11		Kareem	OK
83	LR2	LR2	HIGH STD	10/23/24 21:16		Kareem	OK
84	CCV07	CCV07	CCV	10/23/24 21:28		Kareem	OK
85	CCB07	CCB07	CCB	10/23/24 21:34		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133086

Review By	kareem	Review On	10/25/2024 3:10:26 PM
Supervise By	jaswal	Supervise On	10/26/2024 8:07:20 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441 MP82476 MP82477 MP82478 MP82479 MP82712
ICV Standard	mp82485
CCV Standard	MP82488
ICSA Standard	MP82486 MP82487
CRI Standard	MP82712
LCS Standard	
Chk Standard	MP82491 MP82492

86	P4397-04ADL	WB-301-SWADL	PS	10/23/24 21:39	Not Required	Kareem	Not Ok
87	P4347-01	EFFLUENT-DAY-1-M	SAM	10/23/24 21:43		Kareem	OK
88	P4347-01DUP	EFFLUENT-DAY-1-M	DUP	10/23/24 21:47		Kareem	OK
89	P4347-01L	EFFLUENT-DAY-1-M	SD	10/23/24 21:52		Kareem	OK
90	P4347-01MS	EFFLUENT-DAY-1-M	MS	10/23/24 21:56	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
91	P4347-01MSD	EFFLUENT-DAY-1-M	MSD	10/23/24 22:04	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
92	P4347-01A	EFFLUENT-DAY-1-M	PS	10/23/24 22:08	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
93	PB164302BL	PB164302BL	MB	10/23/24 22:12		Kareem	OK
94	PB164302BS	PB164302BS	LCS	10/23/24 22:16	0.1 ML OF M6010 AND M6001 WERE ADDED TO 10ML OF THE SAMPLE	Kareem	OK
95	CCV08	CCV08	CCV	10/23/24 22:20		Kareem	OK
96	CCB08	CCB08	CCB	10/23/24 22:25		Kareem	OK