

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV44	Mercury	3.94	4.0	99	90 - 110	CV	10/18/2024	14:30	LB133003

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV42	Mercury	5.15	5.0	103	90 - 110	CV	10/18/2024	14:34	LB133003
CCV43	Mercury	5.09	5.0	102	90 - 110	CV	10/18/2024	15:09	LB133003
CCV44	Mercury	5.31	5.0	106	90 - 110	CV	10/18/2024	15:29	LB133003



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	1000	1000	100	90 - 110	P	10/21/2024	18:38	LB133039

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.5	12.0	96	80 - 120	P	10/21/2024	18:42	LB133039

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4980	5000	100	90 - 110	P	10/21/2024	19:04	LB133039
CCV02	Lead	5120	5000	102	90 - 110	P	10/21/2024	19:59	LB133039
CCV03	Lead	5130	5000	103	90 - 110	P	10/21/2024	20:52	LB133039
CCV04	Lead	5120	5000	102	90 - 110	P	10/21/2024	21:17	LB133039
CCV05	Lead	5220	5000	104	90 - 110	P	10/21/2024	22:09	LB133039
CCV06	Lead	5230	5000	105	90 - 110	P	10/21/2024	22:58	LB133039
CCV07	Lead	5140	5000	103	90 - 110	P	10/21/2024	23:42	LB133039

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	979	1000	98	95 - 105	P	10/22/2024	15:01	LB133061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	12.4	12.0	104	80 - 120	P	10/22/2024	15:16	LB133061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	5020	5000	100	90 - 110	P	10/22/2024	15:52	LB133061
CCV02	Lead	5030	5000	100	90 - 110	P	10/22/2024	16:30	LB133061
CCV03	Lead	5050	5000	101	90 - 110	P	10/22/2024	17:25	LB133061
CCV04	Lead	5000	5000	100	90 - 110	P	10/22/2024	18:19	LB133061
CCV05	Lead	5020	5000	100	90 - 110	P	10/22/2024	19:13	LB133061
CCV06	Lead	5030	5000	101	90 - 110	P	10/22/2024	20:24	LB133061
CCV07	Lead	5000	5000	100	90 - 110	P	10/22/2024	21:16	LB133061
CCV08	Lead	4930	5000	99	90 - 110	P	10/22/2024	22:29	LB133061
CCV09	Lead	5050	5000	101	90 - 110	P	10/22/2024	23:20	LB133061

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	1010	1000	101	95 - 105	P	10/23/2024	14:59	LB133086

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.9	12.0	99	80 - 120	P	10/23/2024	15:11	LB133086

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	5020	5000	100	90 - 110	P	10/23/2024	15:57	LB133086
CCV02	Lead	4810	5000	96	90 - 110	P	10/23/2024	16:48	LB133086
CCV03	Lead	4850	5000	97	90 - 110	P	10/23/2024	17:42	LB133086
CCV04	Lead	5080	5000	102	90 - 110	P	10/23/2024	18:32	LB133086
CCV05	Lead	4910	5000	98	90 - 110	P	10/23/2024	19:23	LB133086
CCV06	Lead	4910	5000	98	90 - 110	P	10/23/2024	20:15	LB133086
CCV07	Lead	4820	5000	96	90 - 110	P	10/23/2024	21:28	LB133086
CCV08	Lead	4980	5000	100	90 - 110	P	10/23/2024	22:20	LB133086
