



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.22	0.2	108	40 - 160	CV	10/18/2024	14:39	LB133003
CRI01	Lead	11.2	12.0	94	40 - 160	P	10/21/2024	18:51	LB133039
CRI01	Lead	11.2	12.0	94	40 - 160	P	10/22/2024	15:24	LB133061
CRI01	Lead	11.1	12.0	92	40 - 160	P	10/23/2024	15:22	LB133086