

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Vermont's Original, LLC **SDG No.:** P4432
Contract: DAIR01 **Lab Code:** CHEM **Case No.:** P4432 **SAS No.:** P4432
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	8.62			-12	12	10/21/2024	18:55	LB133039
ICSAB01	Lead	56.3	49.0	115	37	61	10/21/2024	19:00	LB133039
ICSA01	Lead	8.45			-12	12	10/22/2024	15:29	LB133061
ICSAB01	Lead	56.9	49.0	116	37	61	10/22/2024	15:33	LB133061
ICSA01	Lead	7.85			-12	12	10/23/2024	15:26	LB133086
ICSAB01	Lead	55.0	49.0	112	37	61	10/23/2024	15:32	LB133086